



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB88989

Reagent/Standard	Lot/Log #
conductivity standard	WP54988

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB88989BL		1	21.2	0.92	0.99	07/25/2017	16:05
2	LB88989BS	1413	1	21.5	1237.00	1325.62	07/25/2017	16:10
3	I4346-13		1	22.5	922.00	968.23	07/25/2017	16:15
4	I4354-03		1	19.7	631.00	702.07	07/25/2017	16:20
5	I4354-03DUP		1	19.7	635.00	706.52	07/25/2017	16:25

Result = $\frac{\text{(Reading)}}{1 + 0.0191 \text{ (Temperature - 25)}}$

0388989

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CONDUCTANCE072417

WorkList ID : 101049

Date : 7/24/2017 11:44:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2017	Water	I4354-03	Conductance	Cool 4 deg C	PSEG01	H22	265434-265578-COMP	07/20/2017	SM2510 B

Date/Time 7/25/17 15:30
Received by: AR
Relinquished by: AR

Date/Time 7/25/17 17:10
Received by: AR
Relinquished by: AR