



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB89003

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	07/25/2017	09:35
2	I4362-02		1	>100.0	>212.0	07/25/2017	10:05
3	I4362-02DUP		1	>100.0	>212.0	07/25/2017	10:32
4	I4363-01		1	>100.0	>212.0	07/25/2017	10:55
5	I4363-02		1	>100.0	>212.0	07/25/2017	11:20
6	I4369-08		1	>100.0	>212.0	07/25/2017	11:50
7	I4402-08		1	>100.0	>212.0	07/25/2017	13:10

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP72517

WorkList ID : 101080

Date : 7/25/2017 8:55:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2017	Water	I4362-02	Flash Point	Cool 4 deg C	PSEG01	H41	FRAC-TANK	07/20/2017	1010A
	Water	I4363-01	Flash Point	Cool 4 deg C	DIST01	H32	B-COMB-3	07/20/2017	1010A
	Water	I4363-02	Flash Point	Cool 4 deg C	DIST01	H32	B-COMB-4	07/20/2017	1010A
	Water	I4369-08	Flash Point	Cool 4 deg C	AECO05		WC-1-20170720	07/20/2017	1010A
07/27/2017	Solid	I4378-04	Flash Point	Cool 4 deg C	PSEG02	H42	DRUM-2	07/20/2017	1010A
07/27/2017	Solid	I4378-06	Flash Point	Cool 4 deg C	PSEG02	H42	DRUM-3	07/20/2017	1010A
07/26/2017	Water	I4402-08	Flash Point	Cool 4 deg C	PSEG01	H53	FRAC-TANK-260783	07/21/2017	1010A

Date/Time 7/25/17 9:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 7/25/17 13:15
 Received by: CP
 Relinquished by: KS