


7470A

LB89015
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33045



A= 0.0000e+000

B= 2.9931e-004

C= 9.3272e-003

Rho= 0.9997504

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	11	0.000	11				
0.2	0.200	0.216	0.016	689	0.0 %	689				
2.5	2.500	2.408	-0.092	8013	0.0 %	8013				
5.0	5.000	5.018	0.018	16733	0.0 %	16733				
7.5	7.500	7.647	0.147	25517	0.0 %	25517				
10.0	10.000	9.900	-0.100	33045	0.0 %	33045				

LB89015 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Jul 2017 11:25:30					
Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7470A	PPB	26 Jul 2017 11:28:00	S	Std
0.2 - 1	50.2	689	-	0.2000	7470A	PPB	26 Jul 2017 11:30:08	S	Std
2.5 - 1	52.5	8013	-	2.5000	7470A	PPB	26 Jul 2017 11:32:17	S	Std
5.0 - 1	55	16733	-	5.0000	7470A	PPB	26 Jul 2017 11:34:25	S	Std
7.5 - 1	57.5	25517	-	7.5000	7470A	PPB	26 Jul 2017 11:36:33	S	Std
10.0 - 1	510	33045	-	10.0000	7470A	PPB	26 Jul 2017 11:38:40	S	Std
ICV79 - 1	ICV79	12247	3.6749	-7470A		PPB	26 Jul 2017 11:41:21	U	SMPL
ICB79 - 1	ICB79	-130	-0.0296	-7470A		PPB	26 Jul 2017 11:43:28	U	SMPL
CCV12 - 1	CCV12	15954	4.7844	-7470A		PPB	26 Jul 2017 11:45:34	U	SMPL
CCB12 - 1	CCB12	-117	-0.0257	-7470A		PPB	26 Jul 2017 11:47:41	U	SMPL
CRA - 1	CRA	695	0.2173	-7470A		PPB	26 Jul 2017 11:49:48	U	SMPL
HighStd - 1	HighStd	33364	9.9953	-7470A		PPB	26 Jul 2017 11:51:55	U	SMPL
ChkStd - 1	ChkStd	22844	6.8467	-7470A		PPB	26 Jul 2017 11:54:02	U	SMPL
PB100953BL - 1	PBW	-217	-0.0556	-7470A		PPB	26 Jul 2017 11:56:11	U	SMPL
PB100953BS - 1	LCSW	12898	3.8698	-7470A		PPB	26 Jul 2017 11:58:20	U	SMPL
PB100953TB - 1	PB100908TB	-119	-0.0263	-7470A		PPB	26 Jul 2017 12:00:28	U	SMPL
I4402-06 - 1	ROLL-OFFS-COMP	-12	0.0057	-7470A		PPB	26 Jul 2017 12:02:36	U	SMPL
I4402-08 - 1	FRAC-TANK-260783	11	0.0126	-7470A		PPB	26 Jul 2017 12:04:44	U	SMPL
I4407-02 - 1	HDD-PILWC3	15	0.0138	-7470A		PPB	26 Jul 2017 12:06:53	U	SMPL
I4414-03 - 1	OR-03-072417	-7	0.0072	-7470A		PPB	26 Jul 2017 12:08:59	U	SMPL
CCV13 - 1	CCV13	15991	4.7955	-7470A		PPB	26 Jul 2017 12:11:06	U	SMPL
CCB13 - 1	CCB13	-114	-0.0248	-7470A		PPB	26 Jul 2017 12:13:13	U	SMPL
I4415-03 - 1	SU-02-072417	-3	0.0084	-7470A		PPB	26 Jul 2017 12:15:20	U	SMPL
I4420-07 - 1	2A-1-2-3-4-5-COMP-(0-10)	-15	0.0048	-7470A		PPB	26 Jul 2017 12:17:27	U	SMPL
I4420-14 - 1	2A-6-7-8-9-10-COMP-(0-10)	15	0.0138	-7470A		PPB	26 Jul 2017 12:19:34	U	SMPL
I4422-07 - 1	VNJ-201-COMP	5	0.0108	-7470A		PPB	26 Jul 2017 12:21:41	U	SMPL
I4422-11 - 1	72-12018-COMP	33	0.0192	-7470A		PPB	26 Jul 2017 12:23:48	U	SMPL
I4422-15 - 1	72-11940-COMP	-9	0.0066	-7470A		PPB	26 Jul 2017 12:25:56	U	SMPL
I4424-03 - 1	AU-06-072417-A	-10	0.0063	-7470A		PPB	26 Jul 2017 12:28:04	U	SMPL
I4424-06 - 1	AU-06-072417-B	-15	0.0048	-7470A		PPB	26 Jul 2017 12:30:12	U	SMPL
I4424-09 - 1	AU-06-072417-C	-7	0.0072	-7470A		PPB	26 Jul 2017 12:32:21	U	SMPL
I4424-12 - 1	AU-06-072417-D	-8	0.0069	-7470A		PPB	26 Jul 2017 12:34:28	U	SMPL
CCV14 - 1	CCV14	16286	4.8838	-7470A		PPB	26 Jul 2017 12:36:34	U	SMPL
CCB14 - 1	CCB14	-99	-0.0203	-7470A		PPB	26 Jul 2017 12:38:41	U	SMPL
I4424-12DUP - 1	AU-06-072417-DDUP	20	0.0153	-7470A		PPB	26 Jul 2017 12:40:49	U	SMPL
I4424-12MS - 1	AU-06-072417-DMS	11188	3.3580	-7470A		PPB	26 Jul 2017 12:42:56	U	SMPL
I4424-12MSD - 1	AU-06-072417-DMSD	11733	3.5211	-7470A		PPB	26 Jul 2017 12:45:03	U	SMPL
PB100954BL - 1	PBW	-124	-0.0278	-7470A		PPB	26 Jul 2017 12:47:10	U	SMPL
PB100954BS - 1	LCSW	12059	3.6186	-7470A		PPB	26 Jul 2017 13:46:14	U	SMPL
I4432-02 - 1	MW-1	-78	-0.0140	-7470A		PPB	26 Jul 2017 13:48:22	U	SMPL
I4432-02DUP - 1	MW-1DUP	20	0.0153	-7470A		PPB	26 Jul 2017 13:50:30	U	SMPL
I4432-03 - 1	I4432-02MS	11915	3.5755	-7470A		PPB	26 Jul 2017 13:52:38	U	SMPL
I4432-04 - 1	I4432-02MSD	12753	3.8264	-7470A		PPB	26 Jul 2017 13:54:47	U	SMPL
I4432-05 - 1	MW-2	-77	-0.0137	-7470A		PPB	26 Jul 2017 13:56:54	U	SMPL
CCV15 - 1	CCV15	16073	4.8201	-7470A		PPB	26 Jul 2017 13:59:01	U	SMPL
CCB15 - 1	CCB15	-123	-0.0275	-7470A		PPB	26 Jul 2017 14:01:08	U	SMPL
I4432-06 - 1	DUP	1	0.0096	-7470A		PPB	26 Jul 2017 14:03:15	U	SMPL
I4432-07 - 1	RINSEBLANK-GW	-18	0.0039	-7470A		PPB	26 Jul 2017 14:05:23	U	SMPL
I4432-08 - 1	MW-3	-4	0.0081	-7470A		PPB	26 Jul 2017 14:07:30	U	SMPL
I4432-09 - 1	MW-4	26	0.0171	-7470A		PPB	26 Jul 2017 14:09:37	U	SMPL
I4432-10 - 1	MW-5	170	0.0602	-7470A		PPB	26 Jul 2017 14:11:45	U	SMPL
I4432-11 - 1	MW-3-OLDWELL	132	0.0488	-7470A		PPB	26 Jul 2017 14:13:52	U	SMPL
I4424-12LX5 - 1		6	0.0111	-7470A		PPB	26 Jul 2017 14:16:01	U	SMPL
I4424-12A - 1		12965	3.8898	-7470A		PPB	26 Jul 2017 14:18:09	U	SMPL
I4432-02LX5 - 1		-66	-0.0104	-7470A		PPB	26 Jul 2017 14:20:17	U	SMPL
I4432-02A - 1		13298	3.9895	-7470A		PPB	26 Jul 2017 14:22:24	U	SMPL
CCV16 - 1	CCV16	16147	4.8422	-7470A		PPB	26 Jul 2017 14:24:32	U	SMPL
CCB16 - 1	CCB16	-79	-0.0143	-7470A		PPB	26 Jul 2017 14:26:39	U	SMPL