



Analytical Summary Report

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB89017

Reagent/Standard	Lot/Log #
conductivity standard	WP54988

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB89017BL		1	21.2	0.92	0.99	07/25/2017	16:05
2	LB89017BS	1413	1	21.5	1237.00	1325.62	07/25/2017	16:10
3	I4346-14		1	22.5	922.00	968.23	07/25/2017	16:15
4	I4346-14DUP		1	22.5	921.00	967.18	07/25/2017	16:20

Result = $\frac{\text{(Reading)}}{1 + 0.0191 \text{ (Temperature - 25)}}$