



Analytical Summary Report

Analysis Method: 9050A

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB89018

Reagent/Standard	Lot/Log #
conductivity standard	WP54988

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB89018BL		1	21.2	0.91	0.98	07/25/2017	16:30
2	LB89018BS	1413	1	21.6	1239.00	1325.05	07/25/2017	16:35
3	I4346-15		1	22.5	920.00	966.13	07/25/2017	16:41

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$