


ISM02.4

LB 89028
INSTRUMENT ID: CV1

Linear

 μ Abs.:

49382



A= 0.0000e+000

B= 2.0265e-004

C= -4.6803e-002

Rho= 0.9999329

Accept=Accepted

slope
 $y = 1.215 \times 10^4 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.042	-0.042	22	0.000	22					
0.2	0.200	0.172	-0.028	1079	0.0 %	1079					-14
2.5	2.500	2.558	0.058	12852	0.0 %	12852					2
5.0	5.000	5.059	0.059	25194	0.0 %	25194					1
7.5	7.500	7.494	-0.006	37211	0.0 %	37211					0
10.0	10.000	9.960	-0.040	49382	0.0 %	49382					0

MB
2/26/18

LB89028
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	22	-		0 ISM02.4	PPB	7/26/2017 15:13	Std
	0.2 S0.2	1079	-		0.2 ISM02.4	PPB	7/26/2017 15:15	Std
	2.5 S2.5	12852	-		2.5 ISM02.4	PPB	7/26/2017 15:18	Std
	5 S5	25194	-		5 ISM02.4	PPB	7/26/2017 15:20	Std
	7.5 S7.5	37211	-		7.5 ISM02.4	PPB	7/26/2017 15:22	Std
	10 S10	49382	-		10 ISM02.4	PPB	7/26/2017 15:25	Std
ICV	ICV	19996	4.0053	-	ISM02.4	PPB	7/26/2017 15:28	SMPL
ICB	ICB	-205	-0.0883	-	ISM02.4	PPB	7/26/2017 15:30	SMPL
CCV017	CCV017	25273	5.0747	-	ISM02.4	PPB	7/26/2017 15:32	SMPL
CCB017	CCB017	-228	-0.093	-	ISM02.4	PPB	7/26/2017 15:34	SMPL
PB100962BL	PBS004	-40	-0.0549	-	ISM02.4	PPB	7/26/2017 15:37	SMPL
I4406-01	MJJ2X0	1020	0.1599	-	ISM02.4	PPB	7/26/2017 15:39	SMPL
I4406-02	MJJ2X1	960	0.1477	-	ISM02.4	PPB	7/26/2017 15:41	SMPL
I4406-03	MJJ2X2	1616	0.2807	-	ISM02.4	PPB	7/26/2017 15:44	SMPL
I4406-04	MJJ2X2D	1900	0.3382	-	ISM02.4	PPB	7/26/2017 15:46	SMPL
I4406-05	MJJ2X2S	13976	2.7854	-	ISM02.4	PPB	7/26/2017 15:48	SMPL
I4406-06	MJJ2X3	853	0.1261	-	ISM02.4	PPB	7/26/2017 15:50	SMPL
I4406-07	MJJ2X4	21278	4.2651	-	ISM02.4	PPB	7/26/2017 15:53	SMPL
I4406-08	MJJ2X5	1532	0.2637	-	ISM02.4	PPB	7/26/2017 15:55	SMPL
I4406-09	MJJ2Z7	393	0.0328	-	ISM02.4	PPB	7/26/2017 15:57	SMPL
I4406-10	MJJ2Z8	565	0.0677	-	ISM02.4	PPB	7/26/2017 15:59	SMPL
I4406-11	MJJ2X6	2240	0.4071	-	ISM02.4	PPB	7/26/2017 16:02	SMPL
I4406-12	MJJ2X7	968	0.1494	-	ISM02.4	PPB	7/26/2017 16:04	SMPL
I4406-13	MJJ2X8	1204	0.1972	-	ISM02.4	PPB	7/26/2017 16:06	SMPL
I4406-14	MJJ2X9	625	0.0799	-	ISM02.4	PPB	7/26/2017 16:09	SMPL
I4406-15	MJJ2Y0	851	0.1256	-	ISM02.4	PPB	7/26/2017 16:11	SMPL
I4406-16	MJJ2Y1	1089	0.1739	-	ISM02.4	PPB	7/26/2017 16:13	SMPL
I4406-17	MJJ2Y2	806	0.1165	-	ISM02.4	PPB	7/26/2017 16:15	SMPL
I4406-18	MJJ2Y3	1089	0.1739	-	ISM02.4	PPB	7/26/2017 16:18	SMPL
I4406-19	MJJ2Y4	778	0.1109	-	ISM02.4	PPB	7/26/2017 16:20	SMPL
CCV018	CCV018	23974	4.8115	-	ISM02.4	PPB	7/26/2017 16:22	SMPL
CCB018	CCB018	-149	-0.077	-	ISM02.4	PPB	7/26/2017 16:25	SMPL
I4406-20	MJJ2Y5	609	0.0766	-	ISM02.4	PPB	7/26/2017 16:27	SMPL
I4406-21	MJJ2Y6	662	0.0873	-	ISM02.4	PPB	7/26/2017 16:29	SMPL
I4406-22	MJJ2Y7	1370	0.2308	-	ISM02.4	PPB	7/26/2017 16:31	SMPL
PB100963BL	PBS004	114	-0.0237	-	ISM02.4	PPB	7/26/2017 16:34	SMPL
I4409-01	MJJ2Y8	793	0.1139	-	ISM02.4	PPB	7/26/2017 16:36	SMPL
I4409-02	MJJ2Y9	582	0.0711	-	ISM02.4	PPB	7/26/2017 16:38	SMPL
I4409-03	MJJ2Z0	802	0.1157	-	ISM02.4	PPB	7/26/2017 16:40	SMPL
I4409-04	MJJ2Z0D	822	0.1198	-	ISM02.4	PPB	7/26/2017 16:43	SMPL
I4409-05	MJJ2Z0S	12549	2.4962	-	ISM02.4	PPB	7/26/2017 16:45	SMPL
I4409-06	MJJ2Z1	605	0.0758	-	ISM02.4	PPB	7/26/2017 16:47	SMPL
I4409-07	MJJ2Z2	648	0.0845	-	ISM02.4	PPB	7/26/2017 16:50	SMPL
I4409-08	MJJ2Z3	642	0.0833	-	ISM02.4	PPB	7/26/2017 16:52	SMPL
I4409-09	MJJ2Z4	612	0.0772	-	ISM02.4	PPB	7/26/2017 16:54	SMPL
I4409-10	MJJ2Z5	653	0.0855	-	ISM02.4	PPB	7/26/2017 16:56	SMPL

LB89028
INSTRUMENT ID : CV1

I4409-11	MJJ2Z6	518	0.0582 -	ISM02.4	PPB	7/26/2017 16:59	SMPL
I4409-12	MJJ2Z9	577	0.0701 -	ISM02.4	PPB	7/26/2017 17:01	SMPL
I4409-13	MJJ300	663	0.0876 -	ISM02.4	PPB	7/26/2017 17:03	SMPL
I4409-14	MJJ301	995	0.1548 -	ISM02.4	PPB	7/26/2017 17:06	SMPL
I4409-15	MJJ306	703	0.0957 -	ISM02.4	PPB	7/26/2017 17:08	SMPL
I4409-16	MJJ307	618	0.0784 -	ISM02.4	PPB	7/26/2017 17:10	SMPL
CCV019	CCV019	23422	4.6996 -	ISM02.4	PPB	7/26/2017 17:12	SMPL
CCB019	CCB019	-114	-0.0699 -	ISM02.4	PPB	7/26/2017 17:15	SMPL