

LB890 29

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

7/26/2017 15:14

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.004	0.0	0.112	
ICB1	0.007	0.0	0.003	
CCV1	1.007	0.0	0.113	
CCB1	0.003	0.0	0.003	
PB100913BL	0.003	0.0	0.003	
PB100913BS	1.005	0.0	0.112	
I4407-01	0.005	0.0	0.003	
I4407-01DUP	0.004	0.0	0.003	
I4407-01MS	1.029	0.0	0.115	
I4407-01MSD	1.033	0.0	0.115	
CCV2	1.029	0.0	0.115	
CCB2	0.002	0.0	0.003	

N 12
Mean 0.511
SD 0.5297
CV% 103.65

CB 89029

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 26 14:19:17 2017

Wed Jul 26 15:13:28 2017

A1c

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.054	mg/l	7/26/2017 14:19:17	
0.1PPM	A	Phenolics-	P	0.1059	mg/l	7/26/2017 14:19:18	
0.25PPM	A	Phenolics-	P	0.2547	mg/l	7/26/2017 14:19:19	
0.50PPM	A	Phenolics-	P	0.4978	mg/l	7/26/2017 14:19:20	
1.0PPM	A	Phenolics-	P	0.976	mg/l	7/26/2017 14:19:21	
2.0PPM	A	Phenolics-	P	2.0116	mg/l	7/26/2017 14:19:22	
ICV1	S	Phenolics-	P	1.0041	mg/l	7/26/2017 14:54:46	
ICB1	S	Phenolics-	P	0.0066	mg/l	7/26/2017 14:54:47	
CCV1	S	Phenolics-	P	1.0075	mg/l	7/26/2017 14:54:48	
CCB1	S	Phenolics-	P	0.0025	mg/l	7/26/2017 14:54:49	
PB100913BL	S	Phenolics-	P	0.0032	mg/l	7/26/2017 14:54:50	
PB100913BS	S	Phenolics-	P	1.0053	mg/l	7/26/2017 14:54:51	
I4407-01	S	Phenolics-	P	0.0051	mg/l	7/26/2017 14:54:52	
I4407-01DUP	S	Phenolics-	P	0.004	mg/l	7/26/2017 14:54:53	
I4407-01MS	S	Phenolics-	P	1.029	mg/l	7/26/2017 15:13:25	
I4407-01MSD	S	Phenolics-	P	1.0335	mg/l	7/26/2017 15:13:26	
CCV2	S	Phenolics-	P	1.0292	mg/l	7/26/2017 15:13:27	
CCB2	S	Phenolics-	P	0.0023	mg/l	7/26/2017 15:13:28	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AA

7/26/2017 14:29


Test Phenolics-

Accepted 7/26/2017 14:29

Factor 9.143
 Bias 0.002

Coeff. of det. 0.999718

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0540	0.0500	
2	Phenol-2	0.014	0.1059	0.1000	
3	Phenol-2	0.030	0.2547	0.2500	
4	Phenol-2	0.057	0.4978	0.5000	
5	Phenol-2	0.109	0.9760	1.0000	
6	Phenol-2	0.222	2.0116	2.0000	