



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89046

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/26/2017	16:10
2	CAL2	1	Water			7.01	07/26/2017	16:11
3	CAL3	1	Water			10.02	07/26/2017	16:12
4	ICV	1	Water			7.00	07/26/2017	16:14
5	CCV1	1	Water			2.02	07/26/2017	16:15
6	I4445-03	1	Solid	20.02	20	8.20	07/26/2017	16:16
7	I4448-03	1	Solid	20.04	20	8.95	07/26/2017	16:18
8	I4449-02	1	Solid	20.02	20	8.01	07/26/2017	16:20
9	I4449-04	1	Solid	20.03	20	8.18	07/26/2017	16:21
10	I4465-03	1	Solid	20.02	20	8.72	07/26/2017	16:23
11	I4465-06	1	Solid	20.05	20	8.94	07/26/2017	16:25
12	I4466-03	1	Solid	20.04	20	8.99	07/26/2017	16:26
13	I4467-03	1	Solid	20.02	20	8.24	07/26/2017	16:27
14	I4472-03	1	Solid	20.02	20	6.26	07/26/2017	16:30
15	I4473-05	1	Solid	20.03	20	9.89	07/26/2017	16:31
16	CCV2	1	Water			12.02	07/26/2017	16:32
17	I4473-05DUP	1	Solid	20.05	20	9.90	07/26/2017	16:33
18	CCV3	1	Water			12.01	07/26/2017	16:35

WORKLIST(Hardcopy Internal Chain)

1389046

WorkList Name : CORROSIONITY I4445

WorkList ID : 101146

Date : 7/26/2017 11:08:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2017	Solid	I4445-03	Corrosivity	Cool 4 deg C	PSEG01	I22	BLC-IP-OILY-ROCKS-SAND	07/25/2017	9045C
07/28/2017	Solid	I4448-03	Corrosivity	Cool 4 deg C	PSEG01	I32	TCM-ROW	07/25/2017	9045C
08/08/2017	Solid	I4449-02	Corrosivity	Cool 4 deg C	ROYF02	I32	AOC1-0024	07/25/2017	9045C
08/08/2017	Solid	I4449-04	Corrosivity	Cool 4 deg C	ROYF02	I32	AOC2-0024	07/25/2017	9045C
07/31/2017	Solid	I4465-03	Corrosivity	Cool 4 deg C	PSEG05	I33	OK-03-072617-A	07/26/2017	9045C
07/31/2017	Solid	I4465-06	Corrosivity	Cool 4 deg C	PSEG05	I33	OK-03-072617-B	07/26/2017	9045C
07/31/2017	Solid	I4466-03	Corrosivity	Cool 4 deg C	PSEG05	I33	HD-02-072617	07/26/2017	9045C
	Solid	I4467-03	Corrosivity	Cool 4 deg C	WSP001	I33	B-4(0-10)COMP	07/26/2017	9045C
07/31/2017	Solid	I4472-03	Corrosivity	Cool 4 deg C	PSEG05	I42	TR-04-072617	07/26/2017	9045C
07/31/2017	Solid	I4473-05	Corrosivity	Cool 4 deg C	PSEG01	I43	VNJ-202-286-280	07/26/2017	9045C

Date/Time 07/26/17 11:08 AM
 Received by: 74
 Relinquished by: CD

Date/Time 07/26/17 6:30 PM
 Received by: CD
 Relinquished by: JP