


7470A

LB89047
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34738



A= 0.0000e+000

B= 2.9918e-004

C= -2.4031e-001

Rho= 0.9995176

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	840	0.000	840				
0.2	0.200	0.212	0.012	1512	0.0 %	1512				
2.5	2.500	2.566	0.066	9381	0.0 %	9381				
5.0	5.000	4.983	-0.017	17460	0.0 %	17460				
7.5	7.500	7.275	-0.225	25120	0.0 %	25120				
10.0	10.000	10.152	0.152	34738	0.0 %	34738				

LB89047 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jul 2017 09:48:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	840	-	0.0000	7470A	PPB	27 Jul 2017 10:16:01	S	Std
0.2 - 1	50.2	1512	-	0.2000	7470A	PPB	27 Jul 2017 10:18:09	S	Std
2.5 - 1	52.5	9381	-	2.5000	7470A	PPB	27 Jul 2017 10:20:18	S	Std
5.0 - 1	55	17460	-	5.0000	7470A	PPB	27 Jul 2017 10:22:26	S	Std
7.5 - 1	57.5	25120	-	7.5000	7470A	PPB	27 Jul 2017 10:24:34	S	Std
10.0 - 1	510	34738	-	10.0000	7470A	PPB	27 Jul 2017 10:29:15	S	Std
ICV80 - 1	ICV80	13496	3.7974	-	7470A	PPB	27 Jul 2017 10:31:48	U	SMPL
ICB80 - 1	ICB80	668	-0.0405	-	7470A	PPB	27 Jul 2017 10:33:55	U	SMPL
CCV22 - 1	CCV22	17292	4.9330	-	7470A	PPB	27 Jul 2017 10:36:01	U	SMPL
CCB22 - 1	CCB22	748	-0.0165	-	7470A	PPB	27 Jul 2017 10:38:08	U	SMPL
CRA - 1	CRA	1631	0.2476	-	7470A	PPB	27 Jul 2017 10:40:15	U	SMPL
HighStd - 1	HighStd	35086	10.2566	-	7470A	PPB	27 Jul 2017 10:42:21	U	SMPL
ChkStd - 1	ChkStd	23360	6.7484	-	7470A	PPB	27 Jul 2017 10:44:28	U	SMPL
PB100995BL - 1	PBW	694	-0.0327	-	7470A	PPB	27 Jul 2017 10:49:03	U	SMPL
PB100995BS - 1	LCSW	12508	3.5018	-	7470A	PPB	27 Jul 2017 10:51:10	U	SMPL
PB100995TB - 1	PB100951TB	332	-0.1410	-	7470A	PPB	27 Jul 2017 10:53:17	U	SMPL
I4434-04 - 1	FTRT-PRB	664	-0.0417	-	7470A	PPB	27 Jul 2017 10:55:25	U	SMPL
I4434-05 - 1	FTRT-PSTB	501	-0.0904	-	7470A	PPB	27 Jul 2017 10:57:34	U	SMPL
I4442-04 - 1	VNJ-225	827	0.0071	-	7470A	PPB	27 Jul 2017 10:59:42	U	SMPL
I4444-03 - 1	EO-02-072517-A	299	-0.1509	-	7470A	PPB	27 Jul 2017 11:04:22	U	SMPL
CCV23 - 1	CCV23	17404	4.9665	-	7470A	PPB	27 Jul 2017 11:06:29	U	SMPL
CCB23 - 1	CCB23	836	0.0098	-	7470A	PPB	27 Jul 2017 11:08:35	U	SMPL
I4444-06 - 1	EO-02-072517-B	568	-0.0704	-	7470A	PPB	27 Jul 2017 11:13:29	U	SMPL
I4444-09 - 1	EO-02-072517-C	339	-0.1389	-	7470A	PPB	27 Jul 2017 11:15:36	U	SMPL
I4444-12 - 1	EO-02-072517-D	430	-0.1117	-	7470A	PPB	27 Jul 2017 11:17:43	U	SMPL
I4444-12DUP - 1	EO-02-072517-DDUP	473	-0.0988	-	7470A	PPB	27 Jul 2017 11:19:50	U	SMPL
I4444-12MS - 1	EO-02-072517-DMS	12852	3.6047	-	7470A	PPB	27 Jul 2017 11:21:57	U	SMPL
I4444-12MSD - 1	EO-02-072517-DMSD	13645	3.8419	-	7470A	PPB	27 Jul 2017 11:24:04	U	SMPL
I4444-12LX5 - 1		849	0.0137	-	7470A	PPB	27 Jul 2017 11:26:12	U	SMPL
I4444-12A - 1		13638	3.8398	-	7470A	PPB	27 Jul 2017 11:28:20	U	SMPL
CCV24 - 1	CCV24	17285	4.9309	-	7470A	PPB	27 Jul 2017 11:30:28	U	SMPL
CCB24 - 1	CCB24	786	-0.0052	-	7470A	PPB	27 Jul 2017 11:32:36	U	SMPL