


7471A

LB 89058
INSTRUMENT ID: CX2

Linear

 μ Abs.:

42148



A= 0.0000e+000

B= 2.3779e-004

C= -4.6796e-002

Rho= 0.9999099

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	266	0.000	266				
0.2	0.200	0.217	0.017	1109	0.0 %	1109				
2.5	2.500	2.479	-0.021	10621	0.0 %	10621				
5.0	5.000	4.925	-0.075	20909	0.0 %	20909				
7.5	7.500	7.587	0.087	32102	0.0 %	32102	0			
10.0	10.000	9.976	-0.024	42148	0.0 %	42148				

LB89058 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Jul 2017 14:23:10

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	266	-	0.00007471A	PPB	27 Jul 2017 14:34:15	S	Std
0.2 - 1	50.2	1109	-	0.20007471A	PPB	27 Jul 2017 14:36:32	S	Std
2.5 - 1	52.5	10621	-	2.50007471A	PPB	27 Jul 2017 14:38:50	S	Std
5.0 - 1	55	20909	-	5.00007471A	PPB	27 Jul 2017 14:41:07	S	Std
7.5 - 1	57.5	32102	-	7.50007471A	PPB	27 Jul 2017 14:43:25	S	Std
10.0 - 1	59	42148	-	10.00007471A	PPB	27 Jul 2017 14:45:41	S	Std
ICV81 - 1	ICV81	15929	3.7410	-7471A	PPB	27 Jul 2017 14:48:27	U	SMPL
ICB81 - 1	ICB81	5	-0.0456	-7471A	PPB	27 Jul 2017 14:50:43	U	SMPL
CCV25 - 1	CCV25	20363	4.7954	-7471A	PPB	27 Jul 2017 14:52:58	U	SMPL
CCB25 - 1	CCB25	2	-0.0463	-7471A	PPB	27 Jul 2017 14:55:14	U	SMPL
CRA - 1	CRA	1087	0.2117	-7471A	PPB	27 Jul 2017 14:57:30	U	SMPL
HighStd - 1	HighStd	42174	9.9819	-7471A	PPB	27 Jul 2017 14:59:46	U	SMPL
ChkStd - 1	ChkStd	28559	6.7444	-7471A	PPB	27 Jul 2017 15:02:02	U	SMPL
PB100996BL - 1	PBS	-78	-0.0653	-7471A	PPB	27 Jul 2017 15:04:20	U	SMPL
PB100996BS - 1	LCSS	13975	3.2764	-7471A	PPB	27 Jul 2017 15:06:38	U	SMPL
I4445-01 - 1	BLC-IP-OILY-ROCKS-SAND	312	0.0274	-7471A	PPB	27 Jul 2017 15:08:54	U	SMPL
I4446-01 - 1	BLC-UG-PC	7228	1.6720	-7471A	PPB	27 Jul 2017 15:11:12	U	SMPL
I4448-01 - 1	TCM-ROW	1000	0.1910	-7471A	PPB	27 Jul 2017 15:13:29	U	SMPL
I4456-01 - 1	TP-7-SS-3	22678	5.3459	-7471A	PPB	27 Jul 2017 15:15:47	U	SMPL
I4456-02 - 1	TP-7D78	19480	4.5854	-7471A	PPB	27 Jul 2017 15:18:03	U	SMPL
CCV26 - 1	CCV26	20124	4.7386	-7471A	PPB	27 Jul 2017 15:20:19	U	SMPL
CCB26 - 1	CCB26	-7	-0.0485	-7471A	PPB	27 Jul 2017 15:22:35	U	SMPL
I4456-03 - 1	TP-8D24	9078	2.1119	-7471A	PPB	27 Jul 2017 15:24:51	U	SMPL
I4456-04 - 1	TP-10D56	2117	0.4566	-7471A	PPB	27 Jul 2017 15:27:08	U	SMPL
I4456-05 - 1	TP-12-SS-4	21472	5.0591	-7471A	PPB	27 Jul 2017 15:29:24	U	SMPL
I4456-06 - 1	DUP	23383	5.5135	-7471A	PPB	27 Jul 2017 15:31:40	U	SMPL
I4456-07 - 1	TP-13-SS-5	14440	3.3869	-7471A	PPB	27 Jul 2017 15:33:57	U	SMPL
I4456-08 - 1	TP-14D24	10922	2.5504	-7471A	PPB	27 Jul 2017 15:36:15	U	SMPL
I4456-09 - 1	TP-15-SS-6	39683	9.3896	-7471A	PPB	27 Jul 2017 15:38:32	U	SMPL
I4456-10 - 1	TP-16-SS-7	8259	1.9171	-7471A	PPB	27 Jul 2017 15:40:50	U	SMPL
I4456-11 - 1	TP-16D24	8729	2.0289	-7471A	PPB	27 Jul 2017 15:43:09	U	SMPL
I4456-12 - 1	TP-17-SS-8	42097	9.9636	-7471A	PPB	27 Jul 2017 15:45:25	U	SMPL
CCV27 - 1	CCV27	20238	4.7657	-7471A	PPB	27 Jul 2017 15:47:42	U	SMPL
CCB27 - 1	CCB27	-9	-0.0489	-7471A	PPB	27 Jul 2017 15:50:00	U	SMPL
I4456-13 - 1	TP-18D24	7327	1.6955	-7471A	PPB	27 Jul 2017 15:52:17	U	SMPL
I4456-13DUP - 1	TP-18D24DUP	7738	1.7933	-7471A	PPB	27 Jul 2017 15:54:33	U	SMPL
I4456-14 - 1	I4456-13MS	25414	5.9965	-7471A	PPB	27 Jul 2017 15:56:49	U	SMPL
I4456-15 - 1	I4456-13MSD	24066	5.6759	-7471A	PPB	27 Jul 2017 16:01:53	U	SMPL
I4465-01 - 1	OK-03-072617-A	1446	0.2971	-7471A	PPB	27 Jul 2017 16:04:11	U	SMPL
I4465-04 - 1	OK-03-072617-B	575	0.0899	-7471A	PPB	27 Jul 2017 16:06:29	U	SMPL
I4466-01 - 1	HD-02-072617	1261	0.2531	-7471A	PPB	27 Jul 2017 16:08:46	U	SMPL
I4467-02 - 1	B-4(0-10)COMP	7373	1.7065	-7471A	PPB	27 Jul 2017 16:11:03	U	SMPL
PB100997BL - 1	PBS	68	-0.0306	-7471A	PPB	27 Jul 2017 16:13:21	U	SMPL
PB100997BS - 1	LCSS	15956	3.7474	-7471A	PPB	27 Jul 2017 16:18:12	U	SMPL
CCV28 - 1	CCV28	20431	4.8116	-7471A	PPB	27 Jul 2017 16:20:27	U	SMPL
CCB28 - 1	CCB28	-42	-0.0568	-7471A	PPB	27 Jul 2017 16:22:44	U	SMPL
I4471-01 - 1	E2-L6-BASE	34406	8.1347	-7471A	PPB	27 Jul 2017 16:25:00	U	SMPL
I4471-02 - 1	E2-L6-WSW	370340	HIGH	-7471A	PPB	27 Jul 2017 16:27:16	X	
I4471-03 - 1	E2-L6-NSW	22323	5.2615	-7471A	PPB	27 Jul 2017 16:34:46	U	SMPL
I4471-03DUP - 1	E2-L6-NSWDUP	22303	5.2567	-7471A	PPB	27 Jul 2017 16:37:02	U	SMPL
I4471-04 - 1	I4471-03MS	38562	9.1230	-7471A	PPB	27 Jul 2017 16:39:19	U	SMPL
I4471-05 - 1	I4471-03MSD	38396	9.0835	-7471A	PPB	27 Jul 2017 16:41:37	U	SMPL
I4471-06 - 1	E2-L6-NSW-DUP	42205	9.9893	-7471A	PPB	27 Jul 2017 16:43:56	U	SMPL
I4471-07 - 1	E2-L6-ESW	103114	24.4731	-7471A	PPB	27 Jul 2017 16:46:15	U	SMPL
I4471-09 - 1	E2-K6-WSW	16031	3.7653	-7471A	PPB	27 Jul 2017 16:48:34	U	SMPL
I4471-13 - 1	E2-L7-BASE	68554	16.2549	-7471A	PPB	27 Jul 2017 16:50:53	U	SMPL
CCV29 - 1	CCV29	20971	4.9400	-7471A	PPB	27 Jul 2017 16:53:09	U	SMPL
CCB29 - 1	CCB29	-97	-0.0699	-7471A	PPB	27 Jul 2017 16:55:28	U	SMPL
I4472-01 - 1	TR-04-072617	4036	0.9129	-7471A	PPB	27 Jul 2017 16:57:45	U	SMPL
I4473-01 - 1	VNJ-202-286-280	476	0.0664	-7471A	PPB	27 Jul 2017 17:00:02	U	SMPL
I4471-02DLX10 - 1		51506	12.2010	-7471A	PPB	27 Jul 2017 17:02:18	U	SMPL
I4456-13LX5 - 1		1286	0.2590	-7471A	PPB	27 Jul 2017 17:04:35	U	SMPL
I4456-13A - 1		23697	5.5882	-7471A	PPB	27 Jul 2017 17:06:53	U	SMPL
I4471-03LX5 - 1		4413	1.0026	-7471A	PPB	27 Jul 2017 17:13:25	U	SMPL
I4471-03A - 1		36245	8.5720	-7471A	PPB	27 Jul 2017 17:19:38	U	SMPL
I4471-02DL2X50 - 1		10170	2.3716	-7471A	PPB	27 Jul 2017 17:25:01	U	SMPL
I4471-07DLX5 - 1		20374	4.7980	-7471A	PPB	27 Jul 2017 17:27:19	U	SMPL
I4471-13DLX5 - 1		13263	3.1071	-7471A	PPB	27 Jul 2017 17:29:36	U	SMPL
CCV30 - 1	CCV30	21189	4.9918	-7471A	PPB	27 Jul 2017 17:34:29	U	SMPL
CCB30 - 1	CCB30	-50	-0.0587	-7471A	PPB	27 Jul 2017 17:36:46	U	SMPL