

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB89096

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP56736
calibration std. phosphate 0.5 ppm	WP56735
calibration std. phosphate 0.3 ppm	WP56734
calibration std. 0 ppm	WP56731
phosphate ICV std.	WP56739
calibration std. phosphate 0.1 ppm	WP56733
calibration std. phosphate 0.05 ppm	WP56732
Combined reagent	WP56738
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate CCV std.	WP56740
Ascorbic Acid	WP56737

Intercept: -0.0011Slope: 0.651Regression: 0.999813

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/28/2017	11:08
2	CAL2	0.05	1	50	50	0.032	07/28/2017	11:08
3	CAL3	0.10	1	50	50	0.065	07/28/2017	11:09
4	CAL4	0.30	1	50	50	0.186	07/28/2017	11:09
5	CAL5	0.50	1	50	50	0.331	07/28/2017	11:10
6	CAL6	1.00	1	50	50	0.649	07/28/2017	11:10



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB89096

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.323	0.498	07/28/2017	11:11
2	ICB		1	50	50	0.000	0.002	07/28/2017	11:11
3	CCV1	0.50	1	50	50	0.330	0.509	07/28/2017	11:12
4	CCB1		1	50	50	0.000	0.002	07/28/2017	11:12
5	PB101050BL		1	50	50	0.000	0.002	07/28/2017	11:13
6	PB101050BS	0.50	1	50	50	0.327	0.504	07/28/2017	11:13
7	I4464-01		1	50	50	0.000	0.002	07/28/2017	11:14
8	I4464-01DUP		1	50	50	0.000	0.002	07/28/2017	11:14
9	I4464-01MS	0.50	1	50	50	0.325	0.501	07/28/2017	11:15
10	I4464-01MSD	0.50	1	50	50	0.321	0.495	07/28/2017	11:15
11	I4464-02		1	50	50	0.000	0.002	07/28/2017	11:16
12	I4481-01		1	50	50	2.382	3.661	07/28/2017	11:16
13	I4481-01		100	50	50	0.186	0.287	07/28/2017	16:17
14	CCV2	0.50	1	50	50	0.328	0.506	07/28/2017	11:17
15	CCB2		1	50	50	0.000	0.002	07/28/2017	11:18