

7471A


LB89098

INSTRUMENT ID: CX1

Linear

 μ Abs.:

39889



A= 0.0000e+000

B= 2.5082e-004

C= -1.7779e-002

Rho= 0.9999950

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	50	0.000	50				
0.2	0.200	0.186	-0.014	812	0.0 %	812				
2.5	2.500	2.518	0.018	10111	0.0 %	10111				
5.0	5.000	5.008	0.008	20037	0.0 %	20037				
7.5	7.500	7.506	0.006	29996	0.0 %	29996	0			
10.0	10.000	9.987	-0.013	39889	0.0 %	39889				

LB89098 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Jul 2017 12:11:27

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	50	-	0.00007471A	PPB	28 Jul 2017 12:26:51	S	Std
0.2 - 1	50.2	812	-	0.20007471A	PPB	28 Jul 2017 12:29:09	S	Std
2.5 - 1	52.5	10111	-	2.50007471A	PPB	28 Jul 2017 12:31:26	S	Std
5.0 - 1	55	20037	-	5.00007471A	PPB	28 Jul 2017 12:33:43	S	Std
7.5 - 1	57.5	29996	-	7.50007471A	PPB	28 Jul 2017 12:36:01	S	Std
10.0 - 1	510	39889	-	10.00007471A	PPB	28 Jul 2017 12:38:18	S	Std
ICV83 - 1	ICV83	15758	3.9347	-7471A	PPB	28 Jul 2017 12:41:02	U	SMPL
ICB83 - 1	ICB83	-114	-0.0464	-7471A	PPB	28 Jul 2017 12:43:17	U	SMPL
CCV34 - 1	CCV34	18425	4.6036	-7471A	PPB	28 Jul 2017 12:45:33	U	SMPL
CCB34 - 1	CCB34	-98	-0.0424	-7471A	PPB	28 Jul 2017 12:47:49	U	SMPL
CRA - 1	CRA	780	0.1779	-7471A	PPB	28 Jul 2017 12:50:04	U	SMPL
HighStd - 1	HighStd	39911	9.9928	-7471A	PPB	28 Jul 2017 12:52:20	U	SMPL
ChkStd - 1	ChkStd	28515	7.1344	-7471A	PPB	28 Jul 2017 12:54:36	U	SMPL
PB101035BL - 1	PBS	-143	-0.0536	-7471A	PPB	28 Jul 2017 12:56:54	U	SMPL
PB101035BS - 1	LCSS	16521	4.1260	-7471A	PPB	28 Jul 2017 12:59:11	U	SMPL
I4349-01 - 1	0522	348174	HIGH	-7471A	PPB	28 Jul 2017 13:01:28	X	
I4476-01 - 1	TP-8-LENOX-AVE	393	0.0808	-7471A	PPB	28 Jul 2017 13:06:06	U	SMPL
I4476-05 - 1	TP-7-LENOX-AVE	722	0.1633	-7471A	PPB	28 Jul 2017 13:08:23	U	SMPL
I4488-01 - 1	CL-02-072717-A	1324	0.3143	-7471A	PPB	28 Jul 2017 13:10:41	U	SMPL
I4488-04 - 1	CL-02-072717-B	1254	0.2968	-7471A	PPB	28 Jul 2017 13:12:57	U	SMPL
CCV35 - 1	CCV35	19111	4.7757	-7471A	PPB	28 Jul 2017 13:15:13	U	SMPL
CCB35 - 1	CCB35	-85	-0.0391	-7471A	PPB	28 Jul 2017 13:17:29	U	SMPL
I4488-07 - 1	CL-02-072717-C	779	0.1776	-7471A	PPB	28 Jul 2017 13:19:45	U	SMPL
I4488-10 - 1	CL-02-072717-D	3678	0.9047	-7471A	PPB	28 Jul 2017 13:22:01	U	SMPL
I4488-13 - 1	CL-02-072717-E	895	0.2067	-7471A	PPB	28 Jul 2017 13:24:17	U	SMPL
I4490-01 - 1	FK-BPM-01-A2-B2-C2-D2	1408	0.3354	-7471A	PPB	28 Jul 2017 13:26:33	U	SMPL
I4490-01DUP - 1	FK-BPM-01-A2-B2-C2-D2DUP	1185	0.2794	-7471A	PPB	28 Jul 2017 13:28:50	U	SMPL
I4490-01MS - 1	FK-BPM-01-A2-B2-C2-D2MS	16035	4.0041	-7471A	PPB	28 Jul 2017 13:31:07	U	SMPL
I4490-01MSD - 1	FK-BPM-01-A2-B2-C2-D2MSD	15656	3.9091	-7471A	PPB	28 Jul 2017 13:33:24	U	SMPL
I4491-01 - 1	FK-BPM-01-C3-C4-C5-C6	1352	0.3213	-7471A	PPB	28 Jul 2017 13:35:42	U	SMPL
I4492-01 - 1	FK-BPM-01-D3-D4-D5-D6	1194	0.2817	-7471A	PPB	28 Jul 2017 13:38:00	U	SMPL
I4493-01 - 1	E2-HS6-NSW	74901	18.7690	-7471A	PPB	28 Jul 2017 13:40:16	U	SMPL
CCV36 - 1	CCV36	18858	4.7122	-7471A	PPB	28 Jul 2017 13:42:32	U	SMPL
CCB36 - 1	CCB36	-139	-0.0526	-7471A	PPB	28 Jul 2017 13:44:51	U	SMPL
I4493-02 - 1	E2-HS6-BASE	31325	7.8392	-7471A	PPB	28 Jul 2017 13:47:08	U	SMPL
I4493-03 - 1	E2-HS6-ESW	99242	24.8743	-7471A	PPB	28 Jul 2017 13:49:24	U	SMPL
I4493-04 - 1	E2-HS6-SSW	53237	13.3352	-7471A	PPB	28 Jul 2017 13:51:41	U	SMPL
I4493-05 - 1	E2-HS6-WSW	48931	12.2552	-7471A	PPB	28 Jul 2017 13:54:00	U	SMPL
I4497-01 - 1	TP-3-4	176	0.0264	-7471A	PPB	28 Jul 2017 13:56:19	U	SMPL
I4497-05 - 1	TP-1	387	0.0793	-7471A	PPB	28 Jul 2017 13:58:38	U	SMPL
I4497-09 - 1	TP-6	547	0.1194	-7471A	PPB	28 Jul 2017 14:00:55	U	SMPL
I4490-01LX5 - 1		114	0.0108	-7471A	PPB	28 Jul 2017 14:11:50	U	SMPL
I4490-01A - 1		16868	4.2131	-7471A	PPB	28 Jul 2017 14:14:06	U	SMPL
CCV37 - 1	CCV37	18908	4.7248	-7471A	PPB	28 Jul 2017 14:16:22	U	SMPL
CCB37 - 1	CCB37	-120	-0.0479	-7471A	PPB	28 Jul 2017 14:18:38	U	SMPL
I4349-01DLX10 - 1	0522	48324	12.1029	-7471A	PPB	28 Jul 2017 14:27:48	U	SMPL
I4493-01DLX5 - 1		14671	3.6620	-7471A	PPB	28 Jul 2017 14:35:13	U	SMPL
I4493-03DLX5 - 1		19255	4.8118	-7471A	PPB	28 Jul 2017 14:37:30	U	SMPL
I4493-04DLX2 - 1		26229	6.5610	-7471A	PPB	28 Jul 2017 14:39:47	U	SMPL
I4493-05DLX2 - 1		23963	5.9927	-7471A	PPB	28 Jul 2017 14:42:04	U	SMPL
CCV38 - 1	CCV38	19177	4.7922	-7471A	PPB	28 Jul 2017 14:44:21	U	SMPL
CCB38 - 1	CCB38	-103	-0.0436	-7471A	PPB	28 Jul 2017 14:46:38	U	SMPL
I4349-01DL2X25 - 1	0522	19538	4.8828	-7471A	PPB	28 Jul 2017 14:55:44	U	SMPL
CCV39 - 1	CCV39	18985	4.7441	-7471A	PPB	28 Jul 2017 14:58:01	U	SMPL
CCB39 - 1	CCB39	-142	-0.0534	-7471A	PPB	28 Jul 2017 15:00:18	U	SMPL