


245.1

LB 89110
INSTRUMENT ID: CY2

Linear

 μ Abs.:

30621



A= 0.0000e+000

B= 3.2786e-004

C= 3.9333e-002

Rho= 0.9998879

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	14	0.000	14				
0.2	0.200	0.230	0.030	582	0.0 %	582				
2.5	2.500	2.473	-0.027	7424	0.0 %	7424				
5.0	5.000	4.920	-0.080	14885	0.0 %	14885				
7.5	7.500	7.454	-0.046	22616	0.0 %	22616				
10.0	10.000	10.079	0.079	30621	0.0 %	30621				

LB89110 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Jul 2017 16:14:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.0000	245.1	PPB	28 Jul 2017 16:19:40	S	Std
0.2 - 1	50.2	582	-	0.2000	245.1	PPB	28 Jul 2017 16:21:48	S	Std
2.5 - 1	52.5	7424	-	2.5000	245.1	PPB	28 Jul 2017 16:23:57	S	Std
5.0 - 1	55	14885	-	5.0000	245.1	PPB	28 Jul 2017 16:26:05	S	Std
7.5 - 1	57.5	22616	-	7.5000	245.1	PPB	28 Jul 2017 16:28:13	S	Std
10.0 - 1	510	30621	-	10.0000	245.1	PPB	28 Jul 2017 16:30:21	S	Std
ICV84 - 1	ICV84	11693	3.8730	-245.1	PPB	PPB	28 Jul 2017 16:33:02	U	SMPL
ICB84 - 1	ICB84	-73	0.0154	-245.1	PPB	PPB	28 Jul 2017 16:35:09	U	SMPL
CCV40 - 1	CCV40	14872	4.9153	-245.1	PPB	PPB	28 Jul 2017 16:37:16	U	SMPL
CCB40 - 1	CCB40	-78	0.0138	-245.1	PPB	PPB	28 Jul 2017 16:39:22	U	SMPL
CRA - 1	CRA	573	0.2272	-245.1	PPB	PPB	28 Jul 2017 16:41:30	U	SMPL
HighStd - 1	HighStd	30823	10.1450	-245.1	PPB	PPB	28 Jul 2017 16:43:37	U	SMPL
ChkStd - 1	ChkStd	21010	6.9277	-245.1	PPB	PPB	28 Jul 2017 16:45:44	U	SMPL
PB101051BL - 1	PBW	-124	-0.0013	-245.1	PPB	PPB	28 Jul 2017 16:47:53	U	SMPL
PB101051BS - 1	LCSW	12540	4.1507	-245.1	PPB	PPB	28 Jul 2017 16:50:02	U	SMPL
I4366-08 - 1	PT-HG-WS	25849	8.5142	-245.1	PPB	PPB	28 Jul 2017 16:56:27	U	SMPL
CCV41 - 1	CCV41	14566	4.8150	-245.1	PPB	PPB	28 Jul 2017 16:58:37	U	SMPL
CCB41 - 1	CCB41	-104	0.0052	-245.1	PPB	PPB	28 Jul 2017 17:00:45	U	SMPL