



Analytical Summary Report

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB89134

Reagent/Standard	Lot/Log #
conductivity standard	WP54988

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB89134BL		1	21.3	0.90	0.97	07/25/2017	16:46
2	LB89134BS	1413	1	21.5	1240.00	1328.83	07/25/2017	16:53
3	I4346-14		1	22.6	923.00	967.34	07/25/2017	16:58

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$