



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB89137

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	07/28/2017	09:30
2	I4346-03		1	75.0	167.0	07/28/2017	10:02
3	I4447-01		1	>100.0	>212.0	07/28/2017	10:35
4	I4447-01DUP		1	>100.0	>212.0	07/28/2017	11:02
5	I4447-04		1	>100.0	>212.0	07/28/2017	11:20

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 89137

WorkList Name : FP72717

WorkList ID : 101190

Date : 7/27/2017 9:01:03 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2017	Solid	I4447-01	Flash Point	Cool 4 deg C	PSEG02	I23	TRANS-1-LIQUID	07/25/2017	1010A
08/01/2017	Solid	I4447-04	Flash Point	Cool 4 deg C	PSEG02	I23	TRANS-2-LIQUID	07/25/2017	1010A

Date/Time 7/28/17 9:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 7/28/17 11:30 AM
 Received by: CP
 Relinquished by: KS