



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB89141
Analysis Date: 07/31/2017
BalanceID: WC SC-3
OvenID: WC Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 07/31/2017
Extraction IN Time: 13:00
Extraction OUT Time: 13:35

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB89141BL	LB89141BL	WATER	2	1000	100	3.1966	3.1966	3.03	3.1966	3.1966	0	0
2	LB89141BS	LB89141BS	WATER	2	1000	100	3.1044	3.1044	3.01	3.1235	3.1235	0.0191	19.1
3	I4464-01	STORMDRAIN-1	WATER	2	1000	100	3.1556	3.1556	3.04	3.1558	3.1558	0.0002	0.2
4	I4464-01DUP	STORMDRAIN-1DUP	WATER	2	1000	100	3.1006	3.1006	3.05	3.1008	3.1008	0.0002	0.2
5	I4464-02	STORMDRAIN-2	WATER	2	1000	100	3.0362	3.0362	3.02	3.0363	3.0363	0.0001	0.1
6	I4464-02MS	STORMDRAIN-2	WATER	2	1000	100	3.1229	3.1229	3.01	3.1416	3.1416	0.0187	18.7



QC Batch# LB89141 Test: TPH Analysis Date: 07/31/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2254
pH Paper 0-14	W2186
Sodium Sulfate	EP1754
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP56784
LCSWD		
MS/MSD	10.00 ML	WP54439

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 15:00
1.0000 gram Balance: 1.0003 In Time: 14:15
Bal Check Time: 13:20

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 16:30
1.0000 gram Balance: 1.0005 Out Time: 16:00
Bal Check Time: 17:00

WORKLIST(Hardcopy Internal Chain)

2389141

WorkList Name : TPH I4509 WorkList ID : 101333 Date : 7/31/2017 12:40:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4464-01-F, E	TPH	Conc H2SO4 4c	KWAR01	I32	STORMDRAIN-1	07/26/2017	1664A
	Water	I4464-02 -F, E	TPH	Conc H2SO4 4c	KWAR01	I32	STORMDRAIN-2	07/26/2017	1664A

Date/Time 07/31/17 2:40 PM
 Received by: JH
 Relinquished by: CA

Date/Time 07/31/17 6:30 PM
 Received by: JH
 Relinquished by: JH