

LB89163


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

33223



A= 0.0000e+000

B= 2.9882e-004

C= 9.0686e-003

Rho= 0.9998630

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	-5	0.000	-5				
0.2	0.200	0.207	0.007	663	0.0 %	663				
2.5	2.500	2.476	-0.024	8257	0.0 %	8257				
5.0	5.000	4.949	-0.051	16532	0.0 %	16532				
7.5	7.500	7.623	0.123	25480	0.0 %	25480				
10.0	10.000	9.937	-0.063	33223	0.0 %	33223				

LB89163 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Aug 2017 10:19:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-5	-	0.0000	7470A	PPB	01 Aug 2017 10:22:47	S	Std
0.2 - 1	50.2	663	-	0.2000	7470A	PPB	01 Aug 2017 10:24:55	S	Std
2.5 - 1	50.5	8257	-	2.5000	7470A	PPB	01 Aug 2017 10:27:04	S	Std
5.0 - 1	55	16532	-	5.0000	7470A	PPB	01 Aug 2017 10:29:12	S	Std
7.5 - 1	57.5	25480	-	7.5000	7470A	PPB	01 Aug 2017 10:31:20	S	Std
10.0 - 1	510	33223	-	10.0000	7470A	PPB	01 Aug 2017 10:33:27	S	Std
ICV88 - 1	ICV88	12867	3.8540	-	7470A	PPB	01 Aug 2017 10:36:01	U	SMPL
ICB88 - 1	ICB88	-112	-0.0244	-	7470A	PPB	01 Aug 2017 10:38:07	U	SMPL
CCV53 - 1	CCV53	16443	4.9225	-	7470A	PPB	01 Aug 2017 10:40:13	U	SMPL
CCB53 - 1	CCB53	-116	-0.0256	-	7470A	PPB	01 Aug 2017 10:42:20	U	SMPL
CRA - 1	CRA	672	0.2099	-	7470A	PPB	01 Aug 2017 10:44:27	U	SMPL
HighStd - 1	HighStd	33162	9.9185	-	7470A	PPB	01 Aug 2017 10:46:34	U	SMPL
ChkStd - 1	ChkStd	23491	7.0286	-	7470A	PPB	01 Aug 2017 10:48:41	U	SMPL
PB101100BL - 1	PBW	-161	-0.0390	-	7470A	PPB	01 Aug 2017 10:50:49	U	SMPL
PB101100BS - 1	LCSW	12577	3.7673	-	7470A	PPB	01 Aug 2017 10:52:58	U	SMPL
PB101100TB - 1	PB101060TB	-86	-0.0166	-	7470A	PPB	01 Aug 2017 10:55:05	U	SMPL
I4497-04 - 1	TP-3-4	-4	0.0079	-	7470A	PPB	01 Aug 2017 10:57:13	U	SMPL
I4497-08 - 1	TP-1	2	0.0097	-	7470A	PPB	01 Aug 2017 10:59:21	U	SMPL
I4497-12 - 1	TP-6	21	0.0153	-	7470A	PPB	01 Aug 2017 11:01:29	U	SMPL
I4511-03 - 1	TEC-PLATT-COMP	9	0.0118	-	7470A	PPB	01 Aug 2017 11:03:36	U	SMPL
CCV54 - 1	CCV54	16555	4.9560	-	7470A	PPB	01 Aug 2017 11:05:43	U	SMPL
CCB54 - 1	CCB54	-87	-0.0169	-	7470A	PPB	01 Aug 2017 11:07:50	U	SMPL
I4512-02 - 1	HDD-PILWC4	-6	0.0073	-	7470A	PPB	01 Aug 2017 11:09:57	U	SMPL
I4524-03 - 1	SU-04-072817	8	0.0115	-	7470A	PPB	01 Aug 2017 11:12:04	U	SMPL
I4524-03DUP - 1	SU-04-072817DUP	5	0.0106	-	7470A	PPB	01 Aug 2017 11:14:11	U	SMPL
I4524-03MS - 1	SU-04-072817MS	12898	3.8632	-	7470A	PPB	01 Aug 2017 11:16:18	U	SMPL
I4524-03MSD - 1	SU-04-072817MSD	12439	3.7261	-	7470A	PPB	01 Aug 2017 11:18:26	U	SMPL
I4524-03LX5 - 1		-101	-0.0211	-	7470A	PPB	01 Aug 2017 11:20:33	U	SMPL
I4524-03A - 1		13694	4.1011	-	7470A	PPB	01 Aug 2017 11:22:42	U	SMPL
CCV55 - 1	CCV55	16631	4.9787	-	7470A	PPB	01 Aug 2017 11:24:50	U	SMPL
CCB55 - 1	CCB55	-108	-0.0232	-	7470A	PPB	01 Aug 2017 11:26:58	U	SMPL