


ISM02.4

LB89164
INSTRUMENT ID: CYL

Linear

 μ Abs.:

43624



A= 0.0000e+000

B= 2.2838e-004

C= -2.8561e-002

Rho= 0.9999374

Accept=Accepted

90%
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.020	-0.020	38	0.000	38					-8
0.2	0.200	0.184	-0.016	930	0.0 %	930					0
2.5	2.500	2.511	0.011	11118	0.0 %	11118					0
5.0	5.000	5.024	0.024	22124	0.0 %	22124					1
7.5	7.500	7.567	0.067	33261	0.0 %	33261					-1
10.0	10.000	9.934	-0.066	43624	0.0 %	43624					

NS
8/11/17

LB89164
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 ISM02.4	PPB	8/1/2017 11:48	Std
	0.2 S0.2	930	-		0.2 ISM02.4	PPB	8/1/2017 11:50	Std
	2.5 S2.5	11118	-		2.5 ISM02.4	PPB	8/1/2017 11:52	Std
	5 S5	22124	-		5 ISM02.4	PPB	8/1/2017 11:54	Std
	7.5 S7.5	33261	-		7.5 ISM02.4	PPB	8/1/2017 11:59	Std
	10 S10	43624	-		10 ISM02.4	PPB	8/1/2017 12:01	Std
ICV	ICV	17320	3.9269 -		ISM02.4	PPB	8/1/2017 12:04	SMPL
ICB	ICB	-134	-0.0592 -		ISM02.4	PPB	8/1/2017 12:06	SMPL
CCV056	CCV056	22163	5.0329 -		ISM02.4	PPB	8/1/2017 12:08	SMPL
CCB056	CCB056	-122	-0.0564 -		ISM02.4	PPB	8/1/2017 12:10	SMPL
PB101134BL	PBS004	-5	-0.0297 -		ISM02.4	PPB	8/1/2017 12:13	SMPL
I4520-01	MBC8B5	1302	0.2688 -		ISM02.4	PPB	8/1/2017 12:15	SMPL
I4520-02	MBC8B7	2666	0.5803 -		ISM02.4	PPB	8/1/2017 12:17	SMPL
I4520-03	MBC8B8	2361	0.5106 -		ISM02.4	PPB	8/1/2017 12:19	SMPL
I4520-04	MBC8B9	967	0.1923 -		ISM02.4	PPB	8/1/2017 12:22	SMPL
I4520-05	MBC8C0	2980	0.652 -		ISM02.4	PPB	8/1/2017 12:24	SMPL
I4520-06	MBC8C0D	2613	0.5682 -		ISM02.4	PPB	8/1/2017 12:26	SMPL
I4520-07	MBC8C0S	13648	3.0883 -		ISM02.4	PPB	8/1/2017 12:29	SMPL
I4520-08	MBC8C1	3528	0.7771 -		ISM02.4	PPB	8/1/2017 12:31	SMPL
CCV057	CCV057	22163	5.0329 -		ISM02.4	PPB	8/1/2017 12:33	SMPL
CCB057	CCB057	-141	-0.0608 -		ISM02.4	PPB	8/1/2017 12:35	SMPL