

7471A


LB 89174

INSTRUMENT ID : CV1

Linear

 μ Abs.:

40370



A= 0.0000e+000

B= 2.4925e-004

C= 7.0471e-005

Rho= 0.9998976

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	13	0.000	13				
0.2	0.200	0.185	-0.015	742	0.0 %	742				
2.5	2.500	2.559	0.059	10267	0.0 %	10267				
5.0	5.000	4.985	-0.015	19998	0.0 %	19998				
7.5	7.500	7.406	-0.094	29712	0.0 %	29712	0			
10.0	10.000	10.062	0.062	40370	0.0 %	40370				

LB89174 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 01 Aug 2017 14:06:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7471A	PPB	01 Aug 2017 14:11:13	S	Std
0.2 - 1	50.2	742	-	0.2000	7471A	PPB	01 Aug 2017 14:13:30	S	Std
2.5 - 1	52.5	10267	-	2.5000	7471A	PPB	01 Aug 2017 14:15:48	S	Std
5.0 - 1	55	19998	-	5.0000	7471A	PPB	01 Aug 2017 14:18:06	S	Std
7.5 - 1	57.5	29712	-	7.5000	7471A	PPB	01 Aug 2017 14:20:23	S	Std
10.0 - 1	310	40370	-	10.0000	7471A	PPB	01 Aug 2017 14:25:13	S	Std
ICV89 - 1	ICV89	15587	3.8851	-	7471A	PPB	01 Aug 2017 14:27:53	U	SMPL
ICB89 - 1	ICB89	-119	-0.0296	-	7471A	PPB	01 Aug 2017 14:30:09	U	SMPL
CCV58 - 1	CCV58	20084	5.0060	-	7471A	PPB	01 Aug 2017 14:32:25	U	SMPL
CCB58 - 1	CCB58	-109	-0.0271	-	7471A	PPB	01 Aug 2017 14:34:40	U	SMPL
CRA - 1	CRA	746	0.1860	-	7471A	PPB	01 Aug 2017 14:36:57	U	SMPL
HighStd - 1	HighStd	40707	10.1463	-	7471A	PPB	01 Aug 2017 14:39:13	U	SMPL
ChkStd - 1	ChkStd	27265	6.7958	-	7471A	PPB	01 Aug 2017 14:41:29	U	SMPL
PB101135BL - 1	PBS	-182	-0.0453	-	7471A	PPB	01 Aug 2017 14:43:47	U	SMPL
PB101135BS - 1	LCSS	14513	3.6174	-	7471A	PPB	01 Aug 2017 14:46:04	U	SMPL
I4534-01 - 1	72-11938	2311	0.5761	-	7471A	PPB	01 Aug 2017 14:48:21	U	SMPL
I4539-01 - 1	TP-2	584	0.1456	-	7471A	PPB	01 Aug 2017 14:50:38	U	SMPL
I4539-05 - 1	TP-5	561	0.1399	-	7471A	PPB	01 Aug 2017 14:52:55	U	SMPL
I4540-01 - 1	72-12003	18717	4.6653	-	7471A	PPB	01 Aug 2017 14:55:13	U	SMPL
I4540-05 - 1	72-12020	9890	2.4651	-	7471A	PPB	01 Aug 2017 14:57:28	U	SMPL
CCV59 - 1	CCV59	19951	4.9728	-	7471A	PPB	01 Aug 2017 14:59:45	U	SMPL
CCB59 - 1	CCB59	-120	-0.0298	-	7471A	PPB	01 Aug 2017 15:02:01	U	SMPL
I4540-09 - 1	72-11948	15461	3.8537	-	7471A	PPB	01 Aug 2017 15:04:16	U	SMPL
I4540-09DUP - 1	72-11948DUP	16850	4.1999	-	7471A	PPB	01 Aug 2017 15:09:10	U	SMPL
I4540-09MS - 1	72-11948MS	31376	7.8205	-	7471A	PPB	01 Aug 2017 15:15:44	U	SMPL
I4540-09MSD - 1	72-11948MSD	30661	7.6423	-	7471A	PPB	01 Aug 2017 15:18:00	U	SMPL
I4540-13 - 1	72-11942	23109	5.7600	-	7471A	PPB	01 Aug 2017 15:20:17	U	SMPL
I4540-17 - 1	VNJ-221	17156	4.2762	-	7471A	PPB	01 Aug 2017 15:22:36	U	SMPL
I4540-21 - 1	348	25636	6.3898	-	7471A	PPB	01 Aug 2017 15:24:53	U	SMPL
I4541-01 - 1	VNJ-220	20294	5.0583	-	7471A	PPB	01 Aug 2017 15:27:11	U	SMPL
I4541-05 - 1	72-11936	35692	8.8963	-	7471A	PPB	01 Aug 2017 15:29:29	U	SMPL
I4541-09 - 1	VNJ-229	22142	5.5189	-	7471A	PPB	01 Aug 2017 15:31:45	U	SMPL
CCV60 - 1	CCV60	19852	4.9482	-	7471A	PPB	01 Aug 2017 15:34:04	U	SMPL
CCB60 - 1	CCB60	-140	-0.0348	-	7471A	PPB	01 Aug 2017 15:36:21	U	SMPL
I4541-13 - 1	72-11944	28582	7.1241	-	7471A	PPB	01 Aug 2017 15:38:38	U	SMPL
I4543-01 - 1	HR-05-073117-A	5691	1.4185	-	7471A	PPB	01 Aug 2017 15:40:54	U	SMPL
I4543-04 - 1	HR-05-073117-B	4888	1.2184	-	7471A	PPB	01 Aug 2017 15:43:11	U	SMPL
I4543-07 - 1	HR-05-073117-C	2786	0.6945	-	7471A	PPB	01 Aug 2017 15:45:28	U	SMPL
I4540-09LX5 - 1		3067	0.7645	-	7471A	PPB	01 Aug 2017 15:47:44	U	SMPL
I4540-09A - 1		30936	7.7108	-	7471A	PPB	01 Aug 2017 15:50:01	U	SMPL
CCV61 - 1	CCV61	20010	4.9875	-	7471A	PPB	01 Aug 2017 15:52:18	U	SMPL
CCB61 - 1	CCB61	-143	-0.0356	-	7471A	PPB	01 Aug 2017 15:54:37	U	SMPL