


7470A

L389191
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33989



A= 0.0000e+000

B= 2.9468e-004

C= 6.5665e-002

Rho= 0.9998969

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.043	0.043	-78	0.000	-78				
0.2	0.200	0.206	0.006	477	0.0 %	477				
2.5	2.500	2.496	-0.004	8247	0.0 %	8247				
5.0	5.000	4.950	-0.050	16576	0.0 %	16576				
7.5	7.500	7.423	-0.077	24969	0.0 %	24969				
10.0	10.000	10.081	0.081	33989	0.0 %	33989				

LB89191 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Aug 2017 08:36:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-78	-	0.0000	7470A	PPB	02 Aug 2017 09:01:12	S	Std
0.2 - 1	SD-2	477	-	0.2000	7470A	PPB	02 Aug 2017 09:03:20	S	Std
2.5 - 1	SD-5	8247	-	2.5000	7470A	PPB	02 Aug 2017 09:05:28	S	Std
5.0 - 1	SD-5	16576	-	5.0000	7470A	PPB	02 Aug 2017 09:07:37	S	Std
7.5 - 1	SD-5	24969	-	7.5000	7470A	PPB	02 Aug 2017 09:09:45	S	Std
10.0 - 1	SD-5	33989	-	10.0000	7470A	PPB	02 Aug 2017 09:11:52	S	Std
ICV90 - 1	ICV90	13575	4.0659	-7470A	PPB	02 Aug 2017 09:14:33	U	SMPL	
ICB90 - 1	ICB90	-181	0.0123	-7470A	PPB	02 Aug 2017 09:16:39	U	SMPL	
CCV62 - 1	CCV62	16615	4.9617	-7470A	PPB	02 Aug 2017 09:18:46	U	SMPL	
CCB62 - 1	CCB62	-243	-0.0059	-7470A	PPB	02 Aug 2017 09:20:52	U	SMPL	
CRA - 1	CRA	482	0.2077	-7470A	PPB	02 Aug 2017 09:23:00	U	SMPL	
HighStd - 1	HighStd	34015	10.0891	-7470A	PPB	02 Aug 2017 09:25:06	U	SMPL	
ChkStd - 1	ChkStd	23156	6.8892	-7470A	PPB	02 Aug 2017 09:27:13	U	SMPL	
PB101162BL - 1	PBW	-260	-0.0110	-7470A	PPB	02 Aug 2017 09:29:21	U	SMPL	
PB101162BS - 1	LCSW	13083	3.9209	-7470A	PPB	02 Aug 2017 09:31:31	U	SMPL	
PB101162TB - 1	PB101162TB	14	0.0698	-7470A	PPB	02 Aug 2017 09:33:38	U	SMPL	
I4539-04 - 1	TP-2	-17	0.0607	-7470A	PPB	02 Aug 2017 09:37:55	U	SMPL	
I4539-08 - 1	TP-5	-75	0.0436	-7470A	PPB	02 Aug 2017 09:40:03	U	SMPL	
I4534-04 - 1	72-11938	-34	0.0556	-7470A	PPB	02 Aug 2017 09:43:25	U	SMPL	
I4540-03 - 1	72-12003	12	0.0692	-7470A	PPB	02 Aug 2017 09:52:33	U	SMPL	
CCV63 - 1	CCV63	16425	4.9058	-7470A	PPB	02 Aug 2017 09:54:40	U	SMPL	
CCB63 - 1	CCB63	0	0.0657	-7470A	PPB	02 Aug 2017 09:56:47	U	SMPL	
I4540-07 - 1	72-12020	74	0.0875	-7470A	PPB	02 Aug 2017 09:58:53	U	SMPL	
I4540-11 - 1	72-11948	-23	0.0589	-7470A	PPB	02 Aug 2017 10:01:00	U	SMPL	
I4540-15 - 1	72-11942	-51	0.0506	-7470A	PPB	02 Aug 2017 10:03:07	U	SMPL	
I4540-19 - 1	VNJ-221	-11	0.0624	-7470A	PPB	02 Aug 2017 10:05:14	U	SMPL	
I4540-23 - 1	348	-33	0.0559	-7470A	PPB	02 Aug 2017 10:07:21	U	SMPL	
I4541-03 - 1	VNJ-220	-48	0.0515	-7470A	PPB	02 Aug 2017 10:09:29	U	SMPL	
I4541-07 - 1	72-11936	0	0.0657	-7470A	PPB	02 Aug 2017 10:11:37	U	SMPL	
I4541-11 - 1	VNJ-229	0	0.0657	-7470A	PPB	02 Aug 2017 10:13:46	U	SMPL	
I4541-15 - 1	72-11944	0	0.0657	-7470A	PPB	02 Aug 2017 10:15:54	U	SMPL	
I4543-03 - 1	HR-05-073117-A	-4	0.0645	-7470A	PPB	02 Aug 2017 10:18:01	U	SMPL	
CCV64 - 1	CCV64	16519	4.9335	-7470A	PPB	02 Aug 2017 10:20:08	U	SMPL	
CCB64 - 1	CCB64	-97	0.0371	-7470A	PPB	02 Aug 2017 10:22:15	U	SMPL	
I4543-06 - 1	HR-05-073117-B	-2	0.0651	-7470A	PPB	02 Aug 2017 10:24:22	U	SMPL	
I4543-09 - 1	HR-05-073117-C	-3	0.0648	-7470A	PPB	02 Aug 2017 10:26:29	U	SMPL	
I4543-09DUP - 1	HR-05-073117-CDUP	-2	0.0651	-7470A	PPB	02 Aug 2017 10:28:35	U	SMPL	
I4543-09MS - 1	HR-05-073117-CMS	13052	3.9118	-7470A	PPB	02 Aug 2017 10:30:42	U	SMPL	
I4543-09MSD - 1	HR-05-073117-CMSD	12525	3.7565	-7470A	PPB	02 Aug 2017 10:32:50	U	SMPL	
PB101166BL - 1	PBW	73	0.0872	-7470A	PPB	02 Aug 2017 10:34:57	U	SMPL	
PB101166BS - 1	LCSW	14432	4.3185	-7470A	PPB	02 Aug 2017 10:37:05	U	SMPL	
I4535-01 - 1	TRC-MW-101I	1231	0.4284	-7470A	PPB	02 Aug 2017 10:41:22	U	SMPL	
I4535-01DUP - 1	TRC-MW-101IDUP	1005	0.3618	-7470A	PPB	02 Aug 2017 10:43:29	U	SMPL	
CCV65 - 1	CCV65	16543	4.9405	-7470A	PPB	02 Aug 2017 10:45:37	U	SMPL	
CCB65 - 1	CCB65	-5	0.0642	-7470A	PPB	02 Aug 2017 10:47:44	U	SMPL	
I4535-01MS - 1	TRC-MW-101IMS	14804	4.4281	-7470A	PPB	02 Aug 2017 10:49:52	U	SMPL	
I4535-01MSD - 1	TRC-MW-101IMSD	13600	4.0733	-7470A	PPB	02 Aug 2017 10:51:59	U	SMPL	
I4535-02 - 1	TRC-MW-101D	13	0.0695	-7470A	PPB	02 Aug 2017 10:54:06	U	SMPL	
I4543-09LX5 - 1		47	0.0795	-7470A	PPB	02 Aug 2017 10:56:13	U	SMPL	
I4543-09A - 1		13883	4.1567	-7470A	PPB	02 Aug 2017 10:58:21	U	SMPL	
I4535-01LX5 - 1		223	0.1314	-7470A	PPB	02 Aug 2017 11:00:29	U	SMPL	
I4535-01A - 1		15331	4.5834	-7470A	PPB	02 Aug 2017 11:02:37	U	SMPL	
I4471-08 - 1	E2-FB-7-26-17	-86	0.0403	-7470A	PPB	02 Aug 2017 11:04:45	U	SMPL	
CCV66 - 1	CCV66	16639	4.9688	-7470A	PPB	02 Aug 2017 11:06:54	U	SMPL	
CCB66 - 1	CCB66	-38	0.0545	-7470A	PPB	02 Aug 2017 11:09:01	U	SMPL	