



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89192

Temp.: 22.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	08/02/2017	10:20
2	CAL2	1	Water			7.01	08/02/2017	10:21
3	CAL3	1	Water			10.02	08/02/2017	10:22
4	ICV	1	Water			7.00	08/02/2017	10:23
5	CCV1	1	Water			2.02	08/02/2017	10:25
6	I4566-01	1	Water			9.27	08/02/2017	10:26
7	I4566-01DUP	1	Water			9.28	08/02/2017	10:27
8	CCV2	1	Water			12.01	08/02/2017	10:28

WORKLIST(Hardcopy Internal Chain)

2389192

WorkList Name : PH I4566

WorkList ID : 101416

Date : 8/2/2017 9:57:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4566-01	pH	Cool 4 deg C	TOTA01	K41	REFUSE-PIT	08/01/2017	SM 4500-PH B

Date/Time 08/02/17 9:00 AM
 Received by: 711
 Relinquished by: CP

Date/Time 08/02/17 11:00 AM
 Received by: CP
 Relinquished by: JD