

7471A


L389193

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40102



A= 0.0000e+000

B= 2.4663e-004

C= -3.6238e-002

Rho= 0.9995464

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	102	0.000	102				
0.2	0.200	0.183	-0.017	889	0.0 %	889				
2.5	2.500	2.513	0.013	10336	0.0 %	10336				
5.0	5.000	4.941	-0.059	20180	0.0 %	20180				
7.5	7.500	7.721	0.221	31452	0.0 %	31452	0			
10.0	10.000	9.854	-0.146	40102	0.0 %	40102				

LB89193 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 02 Aug 2017 11:19:40

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	80	102	-	0.00007471A	PPB	02 Aug 2017 11:21:05	S	Std
0.2 - 1	80.2	889	-	0.20007471A	PPB	02 Aug 2017 11:23:23	S	Std
2.5 - 1	52.5	10336	-	2.50007471A	PPB	02 Aug 2017 11:25:41	S	Std
5.0 - 1	55	20180	-	5.00007471A	PPB	02 Aug 2017 11:27:58	S	Std
7.5 - 1	57.5	31452	-	7.50007471A	PPB	02 Aug 2017 11:30:15	S	Std
10.0 - 1	58	40102	-	10.00007471A	PPB	02 Aug 2017 11:35:07	S	Std
ICV91 - 1	ICV91	15787	3.8572	-7471A	PPB	02 Aug 2017 11:38:13	U	SMPL
ICB91 - 1	ICB91	22	-0.0308	-7471A	PPB	02 Aug 2017 11:40:28	U	SMPL
CCV67 - 1	CCV67	20342	4.9806	-7471A	PPB	02 Aug 2017 11:42:44	U	SMPL
CCB67 - 1	CCB67	18	-0.0318	-7471A	PPB	02 Aug 2017 11:45:00	U	SMPL
CRA - 1	CRA	854	0.1744	-7471A	PPB	02 Aug 2017 11:47:16	U	SMPL
HighStd - 1	HighStd	40563	9.9676	-7471A	PPB	02 Aug 2017 11:49:31	U	SMPL
ChkStd - 1	ChkStd	27797	6.8192	-7471A	PPB	02 Aug 2017 11:51:48	U	SMPL
PB101171BL - 1	PBS	12	-0.0333	-7471A	PPB	02 Aug 2017 11:54:06	U	SMPL
PB101171BS - 1	LCSS	15299	3.7369	-7471A	PPB	02 Aug 2017 11:56:23	U	SMPL
I4542-01 - 1	A4129	389	0.0597	-7471A	PPB	02 Aug 2017 11:58:40	U	SMPL
I4545-01 - 1	E2-O9-BASE	345791	HIGH	-7471A	PPB	02 Aug 2017 12:00:57	X	
I4558-02 - 1	B-6(0-10)COMP	6080	1.4632	-7471A	PPB	02 Aug 2017 12:05:11	U	SMPL
I4558-04 - 1	B-5(0-10)COMP	5056	1.2107	-7471A	PPB	02 Aug 2017 12:07:26	U	SMPL
CCV68 - 1	CCV68	20384	4.9910	-7471A	PPB	02 Aug 2017 12:09:42	U	SMPL
CCB68 - 1	CCB68	25	-0.0301	-7471A	PPB	02 Aug 2017 12:11:58	U	SMPL
I4559-01 - 1	FK-GF-02-008-A	1973	0.4504	-7471A	PPB	02 Aug 2017 12:14:14	U	SMPL
I4559-01DUP - 1	FK-GF-02-008-ADUP	1971	0.4499	-7471A	PPB	02 Aug 2017 12:16:30	U	SMPL
I4559-01MS - 1	FK-GF-02-008-AMS	16569	4.0501	-7471A	PPB	02 Aug 2017 12:18:46	U	SMPL
I4559-01MSD - 1	FK-GF-02-008-AMSD	16011	3.9125	-7471A	PPB	02 Aug 2017 12:21:02	U	SMPL
I4559-03 - 1	FK-GF-02-008-B	2367	0.5475	-7471A	PPB	02 Aug 2017 12:23:19	U	SMPL
I4560-02 - 1	N-STOCKPILE-COMP	3445	0.8134	-7471A	PPB	02 Aug 2017 12:25:36	U	SMPL
I4560-04 - 1	S-STOCKPILE-COMP	6307	1.5192	-7471A	PPB	02 Aug 2017 12:27:53	U	SMPL
I4562-01 - 1	JC-02-080117-A	3702	0.8768	-7471A	PPB	02 Aug 2017 12:30:11	U	SMPL
I4562-04 - 1	JC-02-080117-B	6251	1.5054	-7471A	PPB	02 Aug 2017 12:32:29	U	SMPL
I4562-07 - 1	JC-02-080117-C	3049	0.7157	-7471A	PPB	02 Aug 2017 12:34:45	U	SMPL
CCV69 - 1	CCV69	20270	4.9629	-7471A	PPB	02 Aug 2017 12:37:01	U	SMPL
CCB69 - 1	CCB69	21	-0.0311	-7471A	PPB	02 Aug 2017 12:39:18	U	SMPL
I4562-10 - 1	JC-02-080117-D	3436	0.8112	-7471A	PPB	02 Aug 2017 12:41:34	U	SMPL
I4562-13 - 1	JC-02-080117-E	18974	4.6432	-7471A	PPB	02 Aug 2017 12:43:50	U	SMPL
I4562-16 - 1	JC-02-080117-F	2610	0.6075	-7471A	PPB	02 Aug 2017 12:46:07	U	SMPL
I4562-19 - 1	JC-02-080117-G	3591	0.8494	-7471A	PPB	02 Aug 2017 12:48:23	U	SMPL
I4562-22 - 1	JC-02-080117-H	43094	10.5918	-7471A	PPB	02 Aug 2017 12:50:40	U	SMPL
I4564-01 - 1	E2-O10-ESW	2397	0.5549	-7471A	PPB	02 Aug 2017 12:52:57	U	SMPL
I4559-01LX5 - 1		439	0.0720	-7471A	PPB	02 Aug 2017 12:55:15	U	SMPL
I4559-01A - 1		18015	4.4067	-7471A	PPB	02 Aug 2017 12:57:33	U	SMPL
I4545-01DLX10 - 1		46823	11.5115	-7471A	PPB	02 Aug 2017 12:59:50	U	SMPL
I4546-01 - 1	SCREENED-2-INCH-MINUS-A	477	0.0814	-7471A	PPB	02 Aug 2017 13:02:07	U	SMPL
CCV70 - 1	CCV70	20217	4.9498	-7471A	PPB	02 Aug 2017 13:04:25	U	SMPL
CCB70 - 1	CCB70	-16	-0.0402	-7471A	PPB	02 Aug 2017 13:06:41	U	SMPL
I4562-22DLX2 - 1		21759	5.3301	-7471A	PPB	02 Aug 2017 13:08:57	U	SMPL
I4545-01DLX50 - 1		9141	2.2182	-7471A	PPB	02 Aug 2017 13:11:14	U	SMPL
CCV71 - 1	CCV71	20267	4.9621	-7471A	PPB	02 Aug 2017 13:13:31	U	SMPL
CCB71 - 1	CCB71	7	-0.0345	-7471A	PPB	02 Aug 2017 13:15:47	U	SMPL