

245.1


LB89207

INSTRUMENT ID : CV1

Linear

 μ Abs.:

30603



A= 0.0000e+000

B= 3.2928e-004

C= 9.0327e-003

Rho= 0.9998886

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	63	0.000	63				
0.2	0.200	0.221	0.021	645	0.0 %	645				
2.5	2.500	2.496	-0.004	7554	0.0 %	7554				
5.0	5.000	4.949	-0.051	15001	0.0 %	15001				
7.5	7.500	7.418	-0.082	22499	0.0 %	22499				
10.0	10.000	10.086	0.086	30603	0.0 %	30603				

LB89207 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Aug 2017 17:23:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.0000	245.1	PPB	02 Aug 2017 17:25:56	S	Std
0.2 - 1	50.2	645	-	0.2000	245.1	PPB	02 Aug 2017 17:28:04	S	Std
2.5 - 1	52.5	7554	-	2.5000	245.1	PPB	02 Aug 2017 17:30:13	S	Std
5.0 - 1	55	15001	-	5.0000	245.1	PPB	02 Aug 2017 17:32:21	S	Std
7.5 - 1	57.5	22499	-	7.5000	245.1	PPB	02 Aug 2017 17:34:29	S	Std
10.0 - 1	59.6	30603	-	10.0000	245.1	PPB	02 Aug 2017 17:39:01	S	Std
ICV92 - 1	ICV92	12379	4.0852	-245.1	PPB	PPB	02 Aug 2017 17:47:15	U	SMPL
ICB92 - 1	ICB92	-69	-0.0137	-245.1	PPB	PPB	02 Aug 2017 17:49:22	U	SMPL
CCV72 - 1	CCV72	14814	4.8871	-245.1	PPB	PPB	02 Aug 2017 17:51:28	U	SMPL
CCB72 - 1	CCB72	-61	-0.0111	-245.1	PPB	PPB	02 Aug 2017 17:53:35	U	SMPL
CRA - 1	CRA	619	0.2129	-245.1	PPB	PPB	02 Aug 2017 17:55:42	U	SMPL
HighStd - 1	HighStd	30248	9.9692	-245.1	PPB	PPB	02 Aug 2017 17:57:49	U	SMPL
ChkStd - 1	ChkStd	20549	6.7755	-245.1	PPB	PPB	02 Aug 2017 17:59:56	U	SMPL
PB101206BL - 1	PBW	57	0.0278	-245.1	PPB	PPB	02 Aug 2017 18:02:04	U	SMPL
PB101206BS - 1	LCSW	13140	4.3358	-245.1	PPB	PPB	02 Aug 2017 18:04:13	U	SMPL
I4566-01 - 1	REFUSE-PIT	956	0.3238	-245.1	PPB	PPB	02 Aug 2017 18:16:26	U	SMPL
I4566-01DUP - 1	REFUSE-PITDUP	984	0.3330	-245.1	PPB	PPB	02 Aug 2017 18:18:34	U	SMPL
I4566-01MS - 1	REFUSE-PITMS	13487	4.4501	-245.1	PPB	PPB	02 Aug 2017 18:20:42	U	SMPL
I4566-01MSD - 1	REFUSE-PITMSD	13420	4.4280	-245.1	PPB	PPB	02 Aug 2017 18:22:51	U	SMPL
I4576-01 - 1	COMP	159	0.0614	-245.1	PPB	PPB	02 Aug 2017 18:24:57	U	SMPL
CCV73 - 1	CCV73	14797	4.8815	-245.1	PPB	PPB	02 Aug 2017 18:27:04	U	SMPL
CCB73 - 1	CCB73	-65	-0.0124	-245.1	PPB	PPB	02 Aug 2017 18:29:11	U	SMPL
I4566-01 - 1 Lx5		199	0.0746	-245.1	PPB	PPB	02 Aug 2017 18:31:18	U	SMPL
I4566-01 - 1 A		12348	4.0750	-245.1	PPB	PPB	02 Aug 2017 18:33:25	U	SMPL
CCV74 - 1	CCV74	14584	4.8113	-245.1	PPB	PPB	02 Aug 2017 18:35:32	U	SMPL
CCB74 - 1	CCB74	-66	-0.0127	-245.1	PPB	PPB	02 Aug 2017 18:37:39	U	SMPL