



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89219

Temp.: 22.3 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	08/03/2017	08:00
2	CAL2	1	Water			7.01	08/03/2017	08:01
3	CAL3	1	Water			10.02	08/03/2017	08:02
4	ICV	1	Water			7.00	08/03/2017	08:03
5	CCV1	1	Water			2.01	08/03/2017	08:05
6	I4555-02	1	Water			11.44	08/03/2017	08:07
7	I4555-02DUP	1	Water			11.45	08/03/2017	08:09
8	CCV2	1	Water			12.01	08/03/2017	08:10

2389219

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4555 WorkList ID : 101444 Date : 8/3/2017 7:44:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/02/2017	Water	I4555-02	✓ Corrosivity	Cool 4 deg C	PSEG01	K63	SV21074LM	08/02/2017	9040C

Date/Time 08/03/17 7:44 AM
Received by: JH
Relinquished by:

Date/Time 08/03/17 10:00 AM
Received by: CP
Relinquished by: JH