

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Dhvani

Date: 08/04/2017

DATE & TIME IN: 08/04/2017 08:30

OVEN TEMP: 104 °C

Run Number: LB89220

DATE & TIME OUT: 08/04/2017 12:15

OVEN TEMP: 103 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB89220BL	LB89220BL	59.5039	59.5039	100	59.5039	59.5039	59.5039	0	0
2	LB89220BS	LB89220BS	59.6732	59.6732	100	59.7246	59.7246	59.7246	0.0514	514
3	I4561-01	REACTOR-PORT-1	60.9119	60.9119	5	61.159	61.159	61.159	0.2471	49420
4	I4561-02	REACTOR-PORT-2	60.9112	60.9112	10	61.0836	61.0836	61.0836	0.1724	17240
5	I4561-03	REACTOR-PORT-3	53.6614	53.6614	10	53.787	53.787	53.787	0.1256	12560
6	I4561-04	REACTOR-PORT-4	57.2928	57.2928	10	57.4205	57.4205	57.4205	0.1277	12770
7	I4561-05	REACTOR-PORT-5	61.8623	61.8623	10	61.9793	61.9793	61.9793	0.117	11700
8	I4561-06	REACTOR-PORT-6	60.9358	60.9358	10	61.0548	61.0548	61.0548	0.119	11900
9	I4561-07	REACTOR-PORT-7	54.2563	54.2563	10	54.3671	54.3671	54.3671	0.1108	11080
10	I4561-08	REACTOR-PORT-8	54.1503	54.1503	10	54.2728	54.2728	54.2728	0.1225	12250
11	I4561-09	REACTOR-PORT-9	61.0595	61.0595	10	61.1534	61.1534	61.1534	0.0939	9390
12	I4561-10	REACTOR-PORT-10	58.428	58.428	10	58.5251	58.5251	58.5251	0.0971	9710
13	I4561-11	REACTOR-PORT-11	62.059	62.059	10	62.1452	62.1452	62.1452	0.0862	8620
14	I4561-12	REACTOR-PORT-12	55.9963	55.9963	50	56.0254	56.0254	56.0254	0.0291	582
15	I4561-12DUP	REACTOR-PORT-12DUP	74.5964	74.5964	20	74.6081	74.6081	74.6081	0.0117	585

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS 14561

WorkList ID : 101456

Date : 8/3/2017 9:10:45 AM

LB 89220

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	14561-01	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-1	08/02/2017	SM2540 D
	Water	14561-02	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-2	08/02/2017	SM2540 D
	Water	14561-03	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-3	08/02/2017	SM2540 D
	Water	14561-04	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-4	08/02/2017	SM2540 D
	Water	14561-05	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-5	08/02/2017	SM2540 D
	Water	14561-06	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-6	08/02/2017	SM2540 D
	Water	14561-07	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-7	08/02/2017	SM2540 D
	Water	14561-08	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-8	08/02/2017	SM2540 D
	Water	14561-09	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-9	08/02/2017	SM2540 D
	Water	14561-10	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-10	08/02/2017	SM2540 D
	Water	14561-11	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-11	08/02/2017	SM2540 D
	Water	14561-12	TSS	Cool 4 deg C	M&MM01	K61	REACTOR-PORT-12	08/02/2017	SM2540 D

Date/Time

08/04/17 8:00am

Received by:

PP

Relinquished by:

JS

Date/Time

08/04/17 11:30am

Received by:

JS

Relinquished by:

PP