


7470A

LB 89222
INSTRUMENT ID: C11

Linear

 μ Abs.:

33844



A= 0.0000e+000

B= 2.9403e-004

C= 1.7897e-002

Rho= 0.9999811

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-79	0.000	-79				
0.2	0.200	0.186	-0.014	571	0.0 %	571				
2.5	2.500	2.517	0.017	8499	0.0 %	8499				
5.0	5.000	4.994	-0.006	16924	0.0 %	16924				
7.5	7.500	7.540	0.040	25582	0.0 %	25582				
10.0	10.000	9.969	-0.031	33844	0.0 %	33844				

LB89222 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Aug 2017 09:29:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-79	-	0.00007470A	PPB		03 Aug 2017 09:34:02	S	Std
0.2 - 1	50.2	571	-	0.20007470A	PPB		03 Aug 2017 09:36:11	S	Std
2.5 - 1	52.5	8499	-	2.50007470A	PPB		03 Aug 2017 09:38:19	S	Std
5.0 - 1	55	16924	-	5.00007470A	PPB		03 Aug 2017 09:40:27	S	Std
7.5 - 1	57.5	25582	-	7.50007470A	PPB		03 Aug 2017 09:42:35	S	Std
10.0 - 1	58	33844	-	10.00007470A	PPB		03 Aug 2017 09:44:41	S	Std
ICV93 - 1	ICV93	12863	3.8000	-7470A	PPB		03 Aug 2017 09:47:15	U	SMPL
ICB93 - 1	ICB93	-147	-0.0253	-7470A	PPB		03 Aug 2017 09:49:21	U	SMPL
CCV75 - 1	CCV75	16887	4.9831	-7470A	PPB		03 Aug 2017 09:51:28	U	SMPL
CCB75 - 1	CCB75	-146	-0.0250	-7470A	PPB		03 Aug 2017 09:53:34	U	SMPL
CRA - 1	CRA	713	0.2275	-7470A	PPB		03 Aug 2017 09:55:41	U	SMPL
HighStd - 1	HighStd	34470	10.1530	-7470A	PPB		03 Aug 2017 09:57:48	U	SMPL
ChkStd - 1	ChkStd	23867	7.0355	-7470A	PPB		03 Aug 2017 09:59:55	U	SMPL
PB101207BL - 1	PBW	-152	-0.0268	-7470A	PPB		03 Aug 2017 10:02:03	U	SMPL
PB101207BS - 1	LCSW	12101	3.5759	-7470A	PPB		03 Aug 2017 10:04:11	U	SMPL
PB101207TB - 1	PB101153TB	-148	-0.0256	-7470A	PPB		03 Aug 2017 10:06:19	U	SMPL
I4542-03 - 1	A4129	-2	0.0173	-7470A	PPB		03 Aug 2017 10:08:27	U	SMPL
I4558-05 - 1	B-6(0-10)COMP	-49	0.0035	-7470A	PPB		03 Aug 2017 10:10:35	U	SMPL
I4558-06 - 1	B-5(0-10)COMP	4	0.0191	-7470A	PPB		03 Aug 2017 10:12:44	U	SMPL
I4560-05 - 1	N-STOCKPILE-COMP	22	0.0244	-7470A	PPB		03 Aug 2017 10:14:51	U	SMPL
CCV76 - 1	CCV76	16716	4.9329	-7470A	PPB		03 Aug 2017 10:16:57	U	SMPL
CCB76 - 1	CCB76	-51	0.0029	-7470A	PPB		03 Aug 2017 10:19:04	U	SMPL
I4560-06 - 1	S-STOCKPILE-COMP	106	0.0491	-7470A	PPB		03 Aug 2017 10:21:11	U	SMPL
I4562-03 - 1	JC-02-080117-A	132	0.0567	-7470A	PPB		03 Aug 2017 10:23:18	U	SMPL
I4562-06 - 1	JC-02-080117-B	-8	0.0155	-7470A	PPB		03 Aug 2017 10:25:25	U	SMPL
I4562-09 - 1	JC-02-080117-C	1	0.0182	-7470A	PPB		03 Aug 2017 10:27:32	U	SMPL
I4562-12 - 1	JC-02-080117-D	16	0.0226	-7470A	PPB		03 Aug 2017 10:29:39	U	SMPL
I4562-15 - 1	JC-02-080117-E	-15	0.0135	-7470A	PPB		03 Aug 2017 10:31:47	U	SMPL
I4562-18 - 1	JC-02-080117-F	6	0.0197	-7470A	PPB		03 Aug 2017 10:33:55	U	SMPL
I4562-21 - 1	JC-02-080117-G	7	0.0200	-7470A	PPB		03 Aug 2017 10:36:04	U	SMPL
I4562-24 - 1	JC-02-080117-H	-3	0.0170	-7470A	PPB		03 Aug 2017 10:38:13	U	SMPL
I4562-24DUP - 1	JC-02-080117-HDUP	24	0.0250	-7470A	PPB		03 Aug 2017 10:40:19	U	SMPL
CCV77 - 1	CCV77	16815	4.9620	-7470A	PPB		03 Aug 2017 10:42:26	U	SMPL
CCB77 - 1	CCB77	-77	-0.0047	-7470A	PPB		03 Aug 2017 10:44:33	U	SMPL
I4562-24MS - 1	JC-02-080117-HMS	12225	3.6124	-7470A	PPB		03 Aug 2017 10:46:40	U	SMPL
I4562-24MSD - 1	JC-02-080117-HMSD	12574	3.7150	-7470A	PPB		03 Aug 2017 10:48:48	U	SMPL
I4562-24LX5 - 1		-71	-0.0030	-7470A	PPB		03 Aug 2017 10:50:55	U	SMPL
I4562-24A - 1		13818	4.0808	-7470A	PPB		03 Aug 2017 10:53:02	U	SMPL
CCV78 - 1	CCV78	16931	4.9961	-7470A	PPB		03 Aug 2017 10:55:10	U	SMPL
CCB78 - 1	CCB78	-79	-0.0053	-7470A	PPB		03 Aug 2017 10:57:17	U	SMPL