


ISM02.4

LB 89226
INSTRUMENT ID: CY1

Linear

 μ Abs.:

38097



A= 0.0000e+000

B= 2.6457e-004

C= -3.9529e-002

Rho= 0.9999545

Accept=Accepted

slope
 $y = mx + c$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/L
0.00	0.000	-0.012	-0.012	105	0.000	105					—
0.2	0.200	0.202	0.002	913	0.0 %	913					1
2.5	2.500	2.535	0.035	9730	0.0 %	9730					1
5.0	5.000	5.003	0.003	19058	0.0 %	19058					0
7.5	7.500	7.433	-0.067	28244	0.0 %	28244					-1
10.0	10.000	10.040	0.040	38097	0.0 %	38097					0

NB
8/3/17

LB89226
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	105	-		0 ISM02.4	PPB	8/3/2017 11:08	Std
	0.2 S0.2	913	-		0.2 ISM02.4	PPB	8/3/2017 11:10	Std
	2.5 S2.5	9730	-		2.5 ISM02.4	PPB	8/3/2017 11:12	Std
	5 S5	19058	-		5 ISM02.4	PPB	8/3/2017 11:14	Std
	7.5 S7.5	28244	-		7.5 ISM02.4	PPB	8/3/2017 11:17	Std
	10 S10	38097	-		10 ISM02.4	PPB	8/3/2017 11:19	Std
ICV	ICV	14646	3.8353	-	ISM02.4	PPB	8/3/2017 11:22	SMPL
ICB	ICB	-14	-0.0432	-	ISM02.4	PPB	8/3/2017 11:24	SMPL
CCV079	CCV079	19130	5.0216	-	ISM02.4	PPB	8/3/2017 11:26	SMPL
CCB079	CCB079	-5	-0.0409	-	ISM02.4	PPB	8/3/2017 11:29	SMPL
PB101195BL	PBW010	131	-0.0049	-	ISM02.4	PPB	8/3/2017 11:31	SMPL
PB101195TB	LEB010	120	-0.0078	-	ISM02.4	PPB	8/3/2017 11:33	SMPL
I4412-01	MJJ2Y4	265	0.0306	-	ISM02.4	PPB	8/3/2017 11:35	SMPL
I4412-02	MJJ2Y7	174	0.0065	-	ISM02.4	PPB	8/3/2017 11:38	SMPL
I4412-03	MJJ2Z0	168	0.0049	-	ISM02.4	PPB	8/3/2017 11:40	SMPL
I4412-04	MJJ2Z0D	111	-0.0102	-	ISM02.4	PPB	8/3/2017 11:42	SMPL
I4412-05	MJJ2Z0S	4130	1.0531	-	ISM02.4	PPB	8/3/2017 11:44	SMPL
I4412-06	MJJ2Z3	152	0.0007	-	ISM02.4	PPB	8/3/2017 11:47	SMPL
I4412-07	MJJ2Z6	158	0.0023	-	ISM02.4	PPB	8/3/2017 11:49	SMPL
I4412-08	MJJ2Z9	178	0.0076	-	ISM02.4	PPB	8/3/2017 11:51	SMPL
I4412-09	MJJ306	140	-0.0025	-	ISM02.4	PPB	8/3/2017 11:54	SMPL
PB101196BL	PBW010	136	-0.0035	-	ISM02.4	PPB	8/3/2017 11:56	SMPL
PB101196TB	LEB010	133	-0.0043	-	ISM02.4	PPB	8/3/2017 11:58	SMPL
I4413-01	MJJ2Y7	190	0.0107	-	ISM02.4	PPB	8/3/2017 12:00	SMPL
I4413-02	MJJ2Z0	182	0.0086	-	ISM02.4	PPB	8/3/2017 12:03	SMPL
I4413-03	MJJ2Z0D	191	0.011	-	ISM02.4	PPB	8/3/2017 12:05	SMPL
I4413-04	MJJ2Z0S	4226	1.0785	-	ISM02.4	PPB	8/3/2017 12:07	SMPL
I4413-05	MJJ2Z3	143	-0.0017	-	ISM02.4	PPB	8/3/2017 12:09	SMPL
I4413-06	MJJ2Z6	149	-0.0001	-	ISM02.4	PPB	8/3/2017 12:12	SMPL
I4413-07	MJJ2Z9	159	0.0025	-	ISM02.4	PPB	8/3/2017 12:14	SMPL
CCV080	CCV080	19201	5.0404	-	ISM02.4	PPB	8/3/2017 12:16	SMPL
CCB080	CCB080	12	-0.0364	-	ISM02.4	PPB	8/3/2017 12:19	SMPL
I4413-08	MJJ306	131	-0.0049	-	ISM02.4	PPB	8/3/2017 12:21	SMPL
CCV081	CCV081	19207	5.042	-	ISM02.4	PPB	8/3/2017 12:23	SMPL
CCB081	CCB081	13	-0.0361	-	ISM02.4	PPB	8/3/2017 12:25	SMPL