


ISM02.4

LB89228
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40278



A= 0.0000e+000

B= 2.4771e-004

C= -2.9866e-002

Rho= 0.9999520

Accept=Accepted

*Slope
y = 1/3712 x + 927*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	-0.029	-0.029	3	0.000	3					-12
0.2	0.200	0.177	-0.023	836	0.0 %	836					2
2.5	2.500	2.539	0.039	10369	0.0 %	10369					1
5.0	5.000	5.027	0.027	20415	0.0 %	20415					1
7.5	7.500	7.539	0.039	30553	0.0 %	30553					-1
10.0	10.000	9.948	-0.052	40278	0.0 %	40278					

*MB
8/3/17*

LB89228
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	3	-		0 ISM02.4	PPB	8/3/2017 12:36	Std
	0.2 S0.2	836	-		0.2 ISM02.4	PPB	8/3/2017 12:39	Std
	2.5 S2.5	10369	-		2.5 ISM02.4	PPB	8/3/2017 12:41	Std
	5 S5	20415	-		5 ISM02.4	PPB	8/3/2017 12:43	Std
	7.5 S7.5	30553	-		7.5 ISM02.4	PPB	8/3/2017 12:46	Std
	10 S10	40278	-		10 ISM02.4	PPB	8/3/2017 12:48	Std
ICV	ICV	16013	3.9368 -		ISM02.4	PPB	8/3/2017 12:51	SMPL
ICB	ICB	-154	-0.068 -		ISM02.4	PPB	8/3/2017 12:53	SMPL
CCV082	CCV082	20541	5.0584 -		ISM02.4	PPB	8/3/2017 12:55	SMPL
CCB082	CCB082	-170	-0.072 -		ISM02.4	PPB	8/3/2017 12:57	SMPL
PB101210BL	PBS004	38	-0.0205 -		ISM02.4	PPB	8/3/2017 13:00	SMPL
I4452-01	MH2TR9	481	0.0893 -		ISM02.4	PPB	8/3/2017 13:02	SMPL
I4452-02	MH2TR9D	399	0.069 -		ISM02.4	PPB	8/3/2017 13:04	SMPL
I4452-03	MH2TR9S	10680	2.6157 -		ISM02.4	PPB	8/3/2017 13:07	SMPL
CCV083	CCV083	20415	5.0272 -		ISM02.4	PPB	8/3/2017 13:09	SMPL
CCB083	CCB083	-134	-0.0631 -		ISM02.4	PPB	8/3/2017 13:11	SMPL