


7471A

LB 89234
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38679



A= 0.0000e+000

B= 2.5992e-004

C= 1.1683e-002

Rho= 0.9998682

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	-21	0.000	-21				
0.2	0.200	0.203	0.003	737	0.0 %	737				
2.5	2.500	2.498	-0.002	9566	0.0 %	9566				
5.0	5.000	5.049	0.049	19382	0.0 %	19382				
7.5	7.500	7.378	-0.122	28342	0.0 %	28342	0			
10.0	10.000	10.065	0.065	38679	0.0 %	38679				

LB89234 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 03 Aug 2017 14:33:32

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.00007471A	PPB		03 Aug 2017 14:36:11	S	Std
0.2 - 1	50.2	737	-	0.20007471A	PPB		03 Aug 2017 14:38:28	S	Std
2.5 - 1	52.5	9566	-	2.50007471A	PPB		03 Aug 2017 14:40:46	S	Std
5.0 - 1	55	19382	-	5.00007471A	PPB		03 Aug 2017 14:43:03	S	Std
7.5 - 1	57.5	28342	-	7.50007471A	PPB		03 Aug 2017 14:45:21	S	Std
10.0 - 1	510	38679	-	10.00007471A	PPB		03 Aug 2017 14:47:37	S	Std
ICV94 - 1	ICV94	14780	3.8532	-7471A	PPB		03 Aug 2017 14:50:18	U	SMPL
ICB94 - 1	ICB94	-158	-0.0294	-7471A	PPB		03 Aug 2017 14:52:34	U	SMPL
CCV84 - 1	CCV84	18961	4.9399	-7471A	PPB		03 Aug 2017 14:54:50	U	SMPL
CCB84 - 1	CCB84	-155	-0.0286	-7471A	PPB		03 Aug 2017 14:57:05	U	SMPL
CRA - 1	CRA	733	0.2022	-7471A	PPB		03 Aug 2017 14:59:21	U	SMPL
HighStd - 1	HighStd	38030	9.8963	-7471A	PPB		03 Aug 2017 15:01:37	U	SMPL
ChkStd - 1	ChkStd	26650	6.9384	-7471A	PPB		03 Aug 2017 15:03:53	U	SMPL
PB101212BL - 1	PBS	-194	-0.0387	-7471A	PPB		03 Aug 2017 15:06:11	U	SMPL
PB101212BS - 1	LCSS	14050	3.6635	-7471A	PPB		03 Aug 2017 15:08:29	U	SMPL
I4568-01 - 1	1-N-LOWER-TW	3718	0.9780	-7471A	PPB		03 Aug 2017 15:10:45	U	SMPL
I4568-02 - 1	2-N-UPPER-TW	2041	0.5422	-7471A	PPB		03 Aug 2017 15:13:03	U	SMPL
I4568-03 - 1	3-S-LOWER-TW	3831	1.0074	-7471A	PPB		03 Aug 2017 15:15:20	U	SMPL
I4568-04 - 1	4-S-UPPER-TW	10363	2.7052	-7471A	PPB		03 Aug 2017 15:17:38	U	SMPL
I4582-01 - 1	S1	83167	21.6281	-7471A	PPB		03 Aug 2017 15:19:54	U	SMPL
CCV85 - 1	CCV85	18644	4.8575	-7471A	PPB		03 Aug 2017 15:22:10	U	SMPL
CCB85 - 1	CCB85	-190	-0.0377	-7471A	PPB		03 Aug 2017 15:24:28	U	SMPL
I4582-01DUP - 1	S1DUP	69145	17.9835	-7471A	PPB		03 Aug 2017 15:26:45	U	SMPL
I4582-01MS - 1	S1MS	101558	26.4082	-7471A	PPB		03 Aug 2017 15:34:17	U	SMPL
I4582-01MSD - 1	S1MSD	100137	26.0388	-7471A	PPB		03 Aug 2017 15:36:34	U	SMPL
I4582-02 - 1	S2	35346	9.1986	-7471A	PPB		03 Aug 2017 15:38:53	U	SMPL
I4584-01 - 1	OK-01-080217	272	0.0824	-7471A	PPB		03 Aug 2017 15:41:12	U	SMPL
I4585-01 - 1	OR-01-080217	3276	0.8632	-7471A	PPB		03 Aug 2017 15:43:31	U	SMPL
I4586-01 - 1	HARDWICK-SOIL-COMP	2980	0.7862	-7471A	PPB		03 Aug 2017 15:45:48	U	SMPL
I4588-01 - 1	E2-O10-BASE	103203	26.8357	-7471A	PPB		03 Aug 2017 15:48:06	U	SMPL
I4588-02 - 1	E2-M8-ESW	310481	80.7104	-7471A	PPB		03 Aug 2017 15:50:24	U	SMPL
I4588-03 - 1	E2-N9-BASE	310439	80.6995	-7471A	PPB		03 Aug 2017 15:52:42	U	SMPL
CCV86 - 1	CCV86	17921	4.6696	-7471A	PPB		03 Aug 2017 15:55:21	U	SMPL
CCB86 - 1	CCB86	-415	-0.0962	-7471A	PPB		03 Aug 2017 15:58:02	U	SMPL
I4588-04 - 1	E2-N9-ESW	195856	50.9176	-7471A	PPB		03 Aug 2017 16:00:20	U	SMPL
I4588-05 - 1	E2-CONCRETE-BATCH-16	5164	1.3539	-7471A	PPB		03 Aug 2017 16:02:36	U	SMPL
I4588-06 - 1	E2-CONCRETE-BATCH-17	9234	2.4117	-7471A	PPB		03 Aug 2017 16:05:01	U	SMPL
I4591-01 - 1	DRUM-1-2-COMP	4001	1.0516	-7471A	PPB		03 Aug 2017 16:07:17	U	SMPL
I4591-04 - 1	GRAVEL-PILE	34482	8.9741	-7471A	PPB		03 Aug 2017 16:09:34	U	SMPL
I4592-01 - 1	15-50-RT3460-RT4215-RT3451	1148	0.3101	-7471A	PPB		03 Aug 2017 16:11:51	U	SMPL
I4594-01 - 1	72-12000-R4	395	0.1143	-7471A	PPB		03 Aug 2017 16:14:09	U	SMPL
I4582-01LX5 - 1		16263	4.2387	-7471A	PPB		03 Aug 2017 16:16:27	U	SMPL
I4582-01A - 1		98609	25.6417	-7471A	PPB		03 Aug 2017 16:18:45	U	SMPL
CCV87 - 1	CCV87	18661	4.8620	-7471A	PPB		03 Aug 2017 16:21:01	U	SMPL
CCB87 - 1	CCB87	-174	-0.0335	-7471A	PPB		03 Aug 2017 16:23:20	U	SMPL
I4582-01DLX5 - 1		16509	4.3026	-7471A	PPB		03 Aug 2017 16:25:37	U	SMPL
I4582-01DUPDLX5 - 1		13522	3.5263	-7471A	PPB		03 Aug 2017 16:27:54	U	SMPL
I4582-01MSDLX5 - 1		20483	5.3355	-7471A	PPB		03 Aug 2017 16:30:10	U	SMPL
I4582-01MSDDLX5 - 1		20741	5.4026	-7471A	PPB		03 Aug 2017 16:32:27	U	SMPL
I4582-01LDLX25 - 1		3098	0.8169	-7471A	PPB		03 Aug 2017 16:34:43	U	SMPL
I4582-01ADLX5 - 1		19096	4.9750	-7471A	PPB		03 Aug 2017 16:37:01	U	SMPL
I4588-01DLX5 - 1		20077	5.2300	-7471A	PPB		03 Aug 2017 16:39:17	U	SMPL
I4588-02DLX10 - 1		32947	8.5751	-7471A	PPB		03 Aug 2017 16:41:35	U	SMPL
I4588-03DLX10 - 1		32983	8.5845	-7471A	PPB		03 Aug 2017 16:43:53	U	SMPL
I4588-04DLX10 - 1		20007	5.2118	-7471A	PPB		03 Aug 2017 16:46:12	U	SMPL
CCV88 - 1	CCV88	19162	4.9922	-7471A	PPB		03 Aug 2017 16:48:29	U	SMPL
CCB88 - 1	CCB88	-149	-0.0270	-7471A	PPB		03 Aug 2017 16:50:46	U	SMPL