



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89240

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP56954
Turbidity std, 40NTU	WP56948
turbidity std 5 NTU	WP56951
turbidity std 20 NTU	WP56949
Turbidity std 1 NTU	WP56952
turbidity std 0 NTU	WP56953
turbidity std 10 NTU	WP56950

Intercept: -0.0327

Slope: 0.9763

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.121	08/04/2017	10:30
2	CAL2	1	1	1.080	08/04/2017	10:34
3	CAL3	5	1	4.600	08/04/2017	10:39
4	CAL4	10	1	9.700	08/04/2017	10:44
5	CAL5	20	1	19.400	08/04/2017	10:50
6	CAL6	40	1	39.100	08/04/2017	10:56



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89240

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.320	9.580	08/04/2017	11:01
2	ICB		1	0.079	0.114	08/04/2017	11:07
3	CCV1	10	1	9.900	10.174	08/04/2017	11:12
4	CCB1		1	0.078	0.113	08/04/2017	11:18
5	LB89240BL		1	0.085	0.121	08/04/2017	11:24
6	I4346-09		1	18.300	18.778	08/04/2017	11:30
7	I4366-01		1	7.150	7.357	08/04/2017	11:37
8	CCV2	10	1	9.600	9.867	08/04/2017	11:42
9	CCB2		1	0.075	0.110	08/04/2017	11:47