



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB89245

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	08/03/2017	12:40
2	I4555-02		1	>100.0	>212.0	08/03/2017	13:05
3	I4555-02DUP		1	>100.0	>212.0	08/03/2017	13:32
4	I4589-01		1	>100.0	>212.0	08/03/2017	13:55

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 89245

WorkList Name : FP8317

WorkList ID : 101473

Date : 8/3/2017 12:09:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2017	Water	I4555-02 B	Flash Point	Cool 4 deg C	PSEG01	K63	SV21074LM	08/02/2017	1010A
08/16/2017	Water	I4589-01 I	Flash Point	1:1 HCl to pH <	DIST01	K62	2AVE-05	08/02/2017	1010A

Date/Time 8/3/17 12:10 PM
 Received by: KS
 Relinquished by: CP

Date/Time 8/3/17 14:00
 Received by: CP
 Relinquished by: KS