



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89250

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	08/03/2017	16:00
2	CAL2	1	Water			7.01	08/03/2017	16:01
3	CAL3	1	Water			10.02	08/03/2017	16:02
4	ICV	1	Water			7.01	08/03/2017	16:03
5	CCV1	1	Water			2.02	08/03/2017	16:05
6	I4610-03	1	Solid	20.02	20	9.41	08/03/2017	16:06
7	I4610-06	1	Solid	20.03	20	7.82	08/03/2017	16:07
8	I4610-09	1	Solid	20.05	20	9.64	08/03/2017	16:10
9	I4610-12	1	Solid	20.01	20	8.88	08/03/2017	16:11
10	I4610-15	1	Solid	20.03	20	8.23	08/03/2017	16:12
11	I4611-03	1	Solid	20.05	20	7.79	08/03/2017	16:13
12	I4611-06	1	Solid	20.02	20	8.55	08/03/2017	16:15
13	I4611-09	1	Solid	20.03	20	7.55	08/03/2017	16:19
14	I4613-03	1	Solid	20.02	20	7.69	08/03/2017	16:20
15	I4613-03DUP	1	Solid	20.03	20	7.70	08/03/2017	16:21
16	CCV2	1	Water			12.02	08/03/2017	16:22

2889250

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4611

WorkList ID : 101484

Date : 8/3/2017 2:43:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/08/2017	Solid	I4610-03	Corrosivity	Cool 4 deg C	PSEG05	K51	PL-02-080317-A	08/03/2017	9045C
08/08/2017	Solid	I4610-06	Corrosivity	Cool 4 deg C	PSEG05	K51	PL-02-080317-B	08/03/2017	9045C
08/08/2017	Solid	I4610-09	Corrosivity	Cool 4 deg C	PSEG05	K51	PL-02-080317-C	08/03/2017	9045C
08/08/2017	Solid	I4610-12	Corrosivity	Cool 4 deg C	PSEG05	K51	PL-02-080317-D	08/03/2017	9045C
08/08/2017	Solid	I4610-15	Corrosivity	Cool 4 deg C	PSEG05	K51	PL-02-080317-E	08/03/2017	9045C
08/08/2017	Solid	I4611-03	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-080317-A	08/03/2017	9045C
08/08/2017	Solid	I4611-06	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-080317-B	08/03/2017	9045C
08/08/2017	Solid	I4611-09	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-080317-C	08/03/2017	9045C
08/08/2017	Solid	I4613-03	Corrosivity	Cool 4 deg C	PSEG05	K32	TR-05-080317	08/03/2017	9045C

Date/Time 08/03/17 2:40 PM
Received by: JR
Relinquished by: CR

Date/Time 08/03/17 2:00 PM
Received by: CR
Relinquished by: JR