


ISM02.4

LB 89251
INSTRUMENT ID: CV1

Linear

 μ Abs.:

3846E



A= 0.0000e+000

B= 2.6275e-004

C= -3.1148e-002

Rho= 0.9998847

Accept=Accepted

slope
y=mx+cept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.00	0.000	0.002	0.002	126	0.000	126					1
0.2	0.200	0.205	0.005	898	0.0 %	898					2
2.5	2.500	2.542	0.042	9792	0.0 %	9792					2
5.0	5.000	4.980	-0.020	19072	0.0 %	19072					0
7.5	7.500	7.396	-0.104	28265	0.0 %	28265					1
10.0	10.000	10.076	0.076	38466	0.0 %	38466					1

NB
8/3/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	126	-		0 ISM02.4	PPB	8/3/2017 19:13	Std
	0.2 S0.2	898	-		0.2 ISM02.4	PPB	8/3/2017 19:15	Std
	2.5 S2.5	9792	-		2.5 ISM02.4	PPB	8/3/2017 19:18	Std
	5 S5	19072	-		5 ISM02.4	PPB	8/3/2017 19:20	Std
	7.5 S7.5	28265	-		7.5 ISM02.4	PPB	8/3/2017 19:22	Std
	10 S10	38466	-		10 ISM02.4	PPB	8/3/2017 19:24	Std
ICV	ICV	14618	3.8098 -		ISM02.4	PPB	8/3/2017 19:27	SMPL
ICB	ICB	4	-0.0301 -		ISM02.4	PPB	8/3/2017 19:29	SMPL
CCV091	CCV091	18922	4.9407 -		ISM02.4	PPB	8/3/2017 19:32	SMPL
CCB091	CCB091	-2	-0.0317 -		ISM02.4	PPB	8/3/2017 19:34	SMPL
PB101251BL	PBW010	-28	-0.0385 -		ISM02.4	PPB	8/3/2017 19:36	SMPL
PB101251TB	LEB010	9	-0.0288 -		ISM02.4	PPB	8/3/2017 19:38	SMPL
I4533-01	MESS44	345	0.0595 -		ISM02.4	PPB	8/3/2017 19:41	SMPL
CCV092	CCV092	18989	4.9583 -		ISM02.4	PPB	8/3/2017 19:43	SMPL
CCB092	CCB092	-3	-0.0319 -		ISM02.4	PPB	8/3/2017 19:45	SMPL