



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB89257
Analysis Date: 08/04/2017
BalanceID: WC SC-3
OvenID: EX Oven-02

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 08/04/2017
Extraction IN Time: 08:30
Extraction OUT Time: 09:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB89257BL	LB89257BL	WATER	2	1000	100	3.7587	3.7587	3.02	3.7587	3.7587	0	0
2	LB89257BS	LB89257BS	WATER	2	1000	100	3.136	3.136	3.04	3.1546	3.1546	0.0186	18.6
3	LB89257BSD	LB89257BSD	WATER	2	1000	100	3.0967	3.0967	3.01	3.1155	3.1155	0.0188	18.8
4	I4589-01	2AVE-05	WATER	2	1000	100	3.0747	3.0747	3.03	3.0757	3.0757	0.001	1



QC Batch# LB89257 Test: Non-Polar Ma Analysis Date: 08/04/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2254
pH Paper 0-14	W2248
Sodium Sulfate	EP1762
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP56784
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0021 In OVEN TEMP : 71 Dessicator Time In : 10:25
1.0000 gram Balance: 1.0004 In Time: 09:45
Bal Check Time: 08:40

After Analysis

0.0020 gram Balance: 0.0019 Out OVEN TEMP: 70 Dessicator Time Out: 12:10
1.0000 gram Balance: 1.0005 Out Time: 11:25
Bal Check Time: 12:40

LB 89257

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR i4369

WorkList ID : 101491

Date : 8/4/2017 8:14:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2017	Water	I4589-01-2	Non-Polar Material	1:1 HCl to pH <	DIST01	K62	2AVE-05	08/02/2017	1664A

Date/Time 08/04/17 8:13 AM
 Received by: JD
 Relinquished by: JD

Date/Time 08/04/17 10:30 AM
 Received by: CP
 Relinquished by: JP