



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89260

Temp.: 18.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/04/2017	10:15
2	CAL2	1	Water			7.00	08/04/2017	10:16
3	CAL3	1	Water			10.03	08/04/2017	10:17
4	ICV	1	Water			6.99	08/04/2017	10:18
5	CCV1	1	Water			2.01	08/04/2017	10:20
6	I4607-01	1	Water			3.87	08/04/2017	10:21
7	I4609-01	1	Water			6.64	08/04/2017	10:22
8	I4609-01DUP	1	Water			6.65	08/04/2017	10:23
9	CCV2	1	Water			12.01	08/04/2017	10:25

LB89260

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4607 WorkList ID : 101524 Date : 8/4/2017 9:58:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4607-01	pH	Cool 4 deg C	PHAR02	K43	32220008-3	08/03/2017	SM 4500-PH B
	Water	I4609-01	pH	Cool 4 deg C	PHAR02	K43	15220032-1	08/03/2017	SM 4500-PH B

Date/Time 08/04/17 10:00 AM
Received by: JP
Relinquished by: JP

Date/Time 08/04/17 11:00 AM
Received by: CS
Relinquished by: JP