

Analytical Summary Report

Analysis Method: 9038

ANALYST: AMANDEEP

Parameter: Sulfate

SUPERVISOR REVIEW BY: apatel

Run Number: LB89262

BalanceID: WC SC-4

Reagent/Standard	Lot/Log #
sulfate ICV 20ppm	WP56964
Sulfate LCS std, 20ppm	WP56966
sulfate 0.0ppm	WP56955
sulfate 5ppm	WP56956
sulfate 10ppm	WP56957
sulfate 20ppm	WP56959
sulfate 30ppm	WP56961
sulfate 40ppm	WP56963
sulfate CCV 20ppm	WP56965
Sulfate std, 15ppm	WP56958
Sulfate std, 35ppm	WP56962
Sulfate std, 25ppm	WP56960
conditioning reagent	WP53743
BARIUM CHLORIDE DIHYDRATE ACS 100G	W2210

Intercept: -0.6621

Slope: 3.3574

Regression: 0.999733

Seq	Lab ID	True Val (mg/l)	DF	MATRIX	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	Water	0.150	0.235	0.085	08/04/2017	11:20
2	CAL2	5	1	Water	0.162	16.200	16.038	08/04/2017	11:26
3	CAL3	10	1	Water	0.198	33.900	33.702	08/04/2017	11:33
4	CAL4	15	1	Water	0.190	49.100	48.910	08/04/2017	11:38
5	CAL5	20	1	Water	0.196	66.100	65.904	08/04/2017	11:45
6	CAL6	25	1	Water	0.201	83.100	82.899	08/04/2017	11:52
7	CAL7	30	1	Water	0.250	99.500	99.250	08/04/2017	11:56
8	CAL8	35	1	Water	0.258	116.000	115.742	08/04/2017	12:04
9	CAL9	40	1	Water	0.261	136.100	135.839	08/04/2017	12:10

Analysis Method: 9038
Parameter: Sulfate
Run Number: LB89262

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	MATRIX	Initial Reading	Final Reading	Difference	Result (mg/L)	AnalDate	Anal Time
1	ICV			20	1	Water	0.199	64.700	64.501	19.409	08/04/2017	12:16
2	ICB				1	Water	0.159	0.215	0.056	0.214	08/04/2017	12:23
3	CCV1			20	1	Water	0.248	65.800	65.552	19.722	08/04/2017	12:30
4	CCB1				1	Water	0.160	0.210	0.050	0.212	08/04/2017	12:36
5	LB89262BL	5.00	50		1	Solid	0.158	0.231	0.073	0.219	08/04/2017	12:41
6	LB89262BS	5.01	50	20	1	Solid	0.192	64.100	63.908	19.232	08/04/2017	12:47
7	I4346-05	5.00	50		1	Solid	35.400	559.000	523.600	156.151	08/04/2017	12:54
8	I4346-05	5.00	50		10	Solid	7.950	66.100	58.150	17.517	08/04/2017	12:59
9	CCV2			20	1	Water	0.196	65.900	65.704	19.767	08/04/2017	13:07
10	CCB2				1	Water	0.158	0.248	0.090	0.224	08/04/2017	13:14