


7471A

LB89278
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39621



A= 0.0000e+000

B= 2.5190e-004

C= 3.3530e-002

Rho= 0.9998129

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	-28	0.000	-28				
0.2	0.200	0.236	0.036	804	0.0 %	804				
2.5	2.500	2.496	-0.004	9774	0.0 %	9774				
5.0	5.000	4.850	-0.150	19122	0.0 %	19122				
7.5	7.500	7.577	0.077	29947	0.0 %	29947	0			
10.0	10.000	10.014	0.014	39621	0.0 %	39621				

LB89278 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 04 Aug 2017 15:21:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.00007471A	-7471A	PPB	04 Aug 2017 15:30:56	S	Std
0.2 - 1	50.2	804	-	0.20007471A	-7471A	PPB	04 Aug 2017 15:33:14	S	Std
2.5 - 1	52.5	9774	-	2.50007471A	-7471A	PPB	04 Aug 2017 15:35:32	S	Std
5.0 - 1	55	19122	-	5.00007471A	-7471A	PPB	04 Aug 2017 15:37:49	S	Std
7.5 - 1	57.5	29947	-	7.50007471A	-7471A	PPB	04 Aug 2017 15:40:07	S	Std
10.0 - 1	510	39621	-	10.00007471A	-7471A	PPB	04 Aug 2017 15:42:24	S	Std
ICV96 - 1	ICV96	14842	3.7723	-7471A	-7471A	PPB	04 Aug 2017 15:45:32	U	SMPL
ICB96 - 1	ICB96	-153	-0.0050	-7471A	-7471A	PPB	04 Aug 2017 15:47:47	U	SMPL
CCV01 - 1	CCV01	19455	4.9343	-7471A	-7471A	PPB	04 Aug 2017 15:50:03	U	SMPL
CCB01 - 1	CCB01	-152	-0.0048	-7471A	-7471A	PPB	04 Aug 2017 15:52:19	U	SMPL
CRA - 1	CRA	788	0.2320	-7471A	-7471A	PPB	04 Aug 2017 15:54:34	U	SMPL
HighStd - 1	HighStd	40339	10.1950	-7471A	-7471A	PPB	04 Aug 2017 15:56:50	U	SMPL
ChkStd - 1	ChkStd	27524	6.9669	-7471A	-7471A	PPB	04 Aug 2017 15:59:06	U	SMPL
PB101258BL - 1	PBS	-217	-0.0211	-7471A	-7471A	PPB	04 Aug 2017 16:01:24	U	SMPL
PB101258BS - 1	LCSS	14326	3.6423	-7471A	-7471A	PPB	04 Aug 2017 16:03:42	U	SMPL
I4555-04 - 1	2013	7972	2.0417	-7471A	-7471A	PPB	04 Aug 2017 16:05:59	U	SMPL
I4555-09 - 1	2043	3160	0.8295	-7471A	-7471A	PPB	04 Aug 2017 16:08:16	U	SMPL
I4587-02 - 1	TP-20	16377	4.1589	-7471A	-7471A	PPB	04 Aug 2017 16:10:33	U	SMPL
I4587-03 - 1	TP-23	14310	3.6383	-7471A	-7471A	PPB	04 Aug 2017 16:12:52	U	SMPL
I4587-04 - 1	TP-25	8096	2.0729	-7471A	-7471A	PPB	04 Aug 2017 16:15:08	U	SMPL
CCV02 - 1	CCV02	19377	4.9146	-7471A	-7471A	PPB	04 Aug 2017 16:17:24	U	SMPL
CCB02 - 1	CCB02	-143	-0.0025	-7471A	-7471A	PPB	04 Aug 2017 16:19:40	U	SMPL
I4587-05 - 1	TP-34	5040	1.3031	-7471A	-7471A	PPB	04 Aug 2017 16:21:55	U	SMPL
I4587-06 - 1	TP-36	5212	1.3464	-7471A	-7471A	PPB	04 Aug 2017 16:24:11	U	SMPL
I4608-01 - 1	CLEAN-SAND-3	326	0.1157	-7471A	-7471A	PPB	04 Aug 2017 16:26:28	U	SMPL
I4608-02 - 1	CLEAN-SAND-4	277	0.1033	-7471A	-7471A	PPB	04 Aug 2017 16:28:44	U	SMPL
I4608-03 - 1	CLEAN-SAND-5	252	0.0970	-7471A	-7471A	PPB	04 Aug 2017 16:31:00	U	SMPL
I4610-01 - 1	PL-02-080317-A	3138	0.8240	-7471A	-7471A	PPB	04 Aug 2017 16:33:17	U	SMPL
I4610-04 - 1	PL-02-080317-B	607	0.1864	-7471A	-7471A	PPB	04 Aug 2017 16:35:34	U	SMPL
I4610-07 - 1	PL-02-080317-C	1601	0.4368	-7471A	-7471A	PPB	04 Aug 2017 16:37:51	U	SMPL
I4610-10 - 1	PL-02-080317-D	792	0.2330	-7471A	-7471A	PPB	04 Aug 2017 16:40:10	U	SMPL
I4610-13 - 1	PL-02-080317-E	1290	0.3585	-7471A	-7471A	PPB	04 Aug 2017 16:42:25	U	SMPL
CCV03 - 1	CCV03	19469	4.9378	-7471A	-7471A	PPB	04 Aug 2017 16:44:42	U	SMPL
CCB03 - 1	CCB03	-131	0.0005	-7471A	-7471A	PPB	04 Aug 2017 16:46:58	U	SMPL
I4611-01 - 1	EO-01-080317-A	1660	0.4517	-7471A	-7471A	PPB	04 Aug 2017 16:49:15	U	SMPL
I4611-04 - 1	EO-01-080317-B	11090	2.8271	-7471A	-7471A	PPB	04 Aug 2017 16:51:31	U	SMPL
I4611-04DUP - 1	EO-01-080317-BDUP	10977	2.7987	-7471A	-7471A	PPB	04 Aug 2017 16:53:47	U	SMPL
I4611-04MS - 1	EO-01-080317-BMS	27229	6.8926	-7471A	-7471A	PPB	04 Aug 2017 16:56:03	U	SMPL
I4611-04MSD - 1	EO-01-080317-BMSD	26248	6.6455	-7471A	-7471A	PPB	04 Aug 2017 16:58:20	U	SMPL
I4611-07 - 1	EO-01-080317-C	1726	0.4683	-7471A	-7471A	PPB	04 Aug 2017 17:00:38	U	SMPL
I4613-01 - 1	TR-05-080317	930	0.2678	-7471A	-7471A	PPB	04 Aug 2017 17:02:56	U	SMPL
PB101259BL - 1	PBS	-26	0.0270	-7471A	-7471A	PPB	04 Aug 2017 17:05:14	U	SMPL
PB101259BS - 1	LCSS	14111	3.5881	-7471A	-7471A	PPB	04 Aug 2017 17:07:31	U	SMPL
I4614-01 - 1	E2-N10-BASE	20300	5.1472	-7471A	-7471A	PPB	04 Aug 2017 17:09:47	U	SMPL
CCV04 - 1	CCV04	19546	4.9572	-7471A	-7471A	PPB	04 Aug 2017 17:12:04	U	SMPL
CCB04 - 1	CCB04	-152	-0.0048	-7471A	-7471A	PPB	04 Aug 2017 17:14:20	U	SMPL
I4614-02 - 1	E2-N10-WSW	88115	22.2299	-7471A	-7471A	PPB	04 Aug 2017 17:23:50	U	SMPL
I4614-02DUP - 1	E2-N10-WSWDUP	87489	22.0722	-7471A	-7471A	PPB	04 Aug 2017 17:26:08	U	SMPL
I4614-02MS - 1	E2-N10-WSWMS	102790	25.9266	-7471A	-7471A	PPB	04 Aug 2017 17:28:28	U	SMPL
I4614-02MSD - 1	E2-N10-WSWMSD	106959	26.9768	-7471A	-7471A	PPB	04 Aug 2017 17:33:27	U	SMPL
I4614-03 - 1	E2-N10-ESW	29198	7.3886	-7471A	-7471A	PPB	04 Aug 2017 17:35:45	U	SMPL
I4614-04 - 1	E2-M10-ESW	65059	16.4221	-7471A	-7471A	PPB	04 Aug 2017 17:38:05	U	SMPL
I4611-04LX5 - 1		1619	0.4414	-7471A	-7471A	PPB	04 Aug 2017 17:40:23	U	SMPL
I4611-04A - 1		26921	6.8150	-7471A	-7471A	PPB	04 Aug 2017 17:42:42	U	SMPL
I4614-02LX5 - 1		16445	4.1761	-7471A	-7471A	PPB	04 Aug 2017 17:45:00	U	SMPL
I4614-02A - 1		101327	25.5581	-7471A	-7471A	PPB	04 Aug 2017 17:47:17	U	SMPL
CCV05 - 1	CCV05	19129	4.8522	-7471A	-7471A	PPB	04 Aug 2017 17:49:33	U	SMPL
CCB05 - 1	CCB05	-206	-0.0184	-7471A	-7471A	PPB	04 Aug 2017 17:51:52	U	SMPL
I4614-02DLX5 - 1		16761	4.2557	-7471A	-7471A	PPB	04 Aug 2017 17:54:10	U	SMPL
I4614-02DUPDLX5 - 1		16641	4.2254	-7471A	-7471A	PPB	04 Aug 2017 17:56:26	U	SMPL
I4614-02MSDLX5 - 1		20049	5.0839	-7471A	-7471A	PPB	04 Aug 2017 17:58:42	U	SMPL
I4614-02MSDDLX5 - 1		20800	5.2731	-7471A	-7471A	PPB	04 Aug 2017 18:00:59	U	SMPL
I4614-02LDLX25 - 1		3128	0.8215	-7471A	-7471A	PPB	04 Aug 2017 18:03:16	U	SMPL
I4614-02ADLX5 - 1		20426	5.1789	-7471A	-7471A	PPB	04 Aug 2017 18:13:39	U	SMPL
I4614-04DLX5 - 1		11921	3.0365	-7471A	-7471A	PPB	04 Aug 2017 18:15:56	U	SMPL
CCV06 - 1	CCV06	19498	4.9451	-7471A	-7471A	PPB	04 Aug 2017 18:18:13	U	SMPL
CCB06 - 1	CCB06	-154	-0.0053	-7471A	-7471A	PPB	04 Aug 2017 18:20:31	U	SMPL