


7470A

LB89296
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33543



A= 0.0000e+000

B= 2.9831e-004

C= -1.3680e-002

Rho= 0.9999796

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	27	0.000	27				
0.2	0.200	0.179	-0.021	645	0.0 %	645				
2.5	2.500	2.547	0.047	8585	0.0 %	8585				
5.0	5.000	4.978	-0.022	16734	0.0 %	16734				
7.5	7.500	7.509	0.009	25218	0.0 %	25218				
10.0	10.000	9.992	-0.008	33543	0.0 %	33543				

LB89296 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Aug 2017 10:12:34

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	27	-	0.00007470A	PPB		07 Aug 2017 10:27:16	S	Std
0.2 - 1	S0.2	645	-	0.20007470A	PPB		07 Aug 2017 10:29:24	S	Std
2.5 - 1	S2.5	8585	-	2.50007470A	PPB		07 Aug 2017 10:31:33	S	Std
5.0 - 1	S5	16734	-	5.00007470A	PPB		07 Aug 2017 10:33:41	S	Std
7.5 - 1	S7.5	25218	-	7.50007470A	PPB		07 Aug 2017 10:35:49	S	Std
10.0 - 1	S10	33543	-	10.00007470A	PPB		07 Aug 2017 10:40:34	S	Std
ICV97 - 1	ICV97	12870	3.8255	-7470A	PPB		07 Aug 2017 10:43:05	U	SMPL
ICB97 - 1	ICB97	-171	-0.0647	-7470A	PPB		07 Aug 2017 10:45:11	U	SMPL
CCV07 - 1	CCV07	16909	5.0304	-7470A	PPB		07 Aug 2017 10:47:18	U	SMPL
CCB07 - 1	CCB07	-167	-0.0635	-7470A	PPB		07 Aug 2017 10:49:24	U	SMPL
CRA - 1	CRA	678	0.1886	-7470A	PPB		07 Aug 2017 10:51:32	U	SMPL
HighStd - 1	HighStd	33639	10.0211	-7470A	PPB		07 Aug 2017 10:53:38	U	SMPL
ChkStd - 1	ChkStd	23435	6.9771	-7470A	PPB		07 Aug 2017 10:55:45	U	SMPL
PB101288BL - 1	PBW	-126	-0.0513	-7470A	PPB		07 Aug 2017 10:57:54	U	SMPL
PB101288BS - 1	LCSW	12754	3.7909	-7470A	PPB		07 Aug 2017 11:00:02	U	SMPL
PB101288TB - 1	PB101252TB	-141	-0.0557	-7470A	PPB		07 Aug 2017 11:02:10	U	SMPL
I4592-02 - 1	15-50-RT3460-RT4215-RT3451	-122	-0.0501	-7470A	PPB		07 Aug 2017 11:04:18	U	SMPL
I4594-05 - 1	72-12000-R4	0	-0.0137	-7470A	PPB		07 Aug 2017 11:06:26	U	SMPL
I4610-03 - 1	PL-02-080317-A	5	-0.0122	-7470A	PPB		07 Aug 2017 11:08:35	U	SMPL
I4610-06 - 1	PL-02-080317-B	43	-0.0009	-7470A	PPB		07 Aug 2017 11:10:42	U	SMPL
CCV08 - 1	CCV08	17196	5.1160	-7470A	PPB		07 Aug 2017 11:12:49	U	SMPL
CCB08 - 1	CCB08	-145	-0.0569	-7470A	PPB		07 Aug 2017 11:14:55	U	SMPL
I4610-09 - 1	PL-02-080317-C	-58	-0.0310	-7470A	PPB		07 Aug 2017 11:17:02	U	SMPL
I4610-12 - 1	PL-02-080317-D	-12	-0.0173	-7470A	PPB		07 Aug 2017 11:19:09	U	SMPL
I4610-15 - 1	PL-02-080317-E	19	-0.0080	-7470A	PPB		07 Aug 2017 11:21:15	U	SMPL
I4611-03 - 1	EO-01-080317-A	26	-0.0059	-7470A	PPB		07 Aug 2017 11:23:23	U	SMPL
I4611-06 - 1	EO-01-080317-B	38	-0.0023	-7470A	PPB		07 Aug 2017 11:25:30	U	SMPL
I4611-09 - 1	EO-01-080317-C	-53	-0.0295	-7470A	PPB		07 Aug 2017 11:27:38	U	SMPL
I4613-03 - 1	TR-05-080317	-49	-0.0283	-7470A	PPB		07 Aug 2017 11:29:46	U	SMPL
I4613-03DUP - 1	TR-05-080317DUP	-27	-0.0217	-7470A	PPB		07 Aug 2017 11:31:54	U	SMPL
I4613-03MS - 1	TR-05-080317MS	12656	3.7617	-7470A	PPB		07 Aug 2017 11:34:03	U	SMPL
I4613-03MSD - 1	TR-05-080317MSD	12392	3.6829	-7470A	PPB		07 Aug 2017 11:36:10	U	SMPL
CCV09 - 1	CCV09	17397	5.1760	-7470A	PPB		07 Aug 2017 11:38:18	U	SMPL
CCB09 - 1	CCB09	-137	-0.0545	-7470A	PPB		07 Aug 2017 11:40:25	U	SMPL
I4613-03LX5 - 1		-58	-0.0310	-7470A	PPB		07 Aug 2017 11:42:31	U	SMPL
I4613-03A - 1		12701	3.7751	-7470A	PPB		07 Aug 2017 11:44:38	U	SMPL
CCV10 - 1	CCV10	17192	5.1148	-7470A	PPB		07 Aug 2017 11:46:45	U	SMPL
CCB10 - 1	CCB10	-87	-0.0396	-7470A	PPB		07 Aug 2017 11:48:52	U	SMPL