

245.1


L989301

INSTRUMENT ID: CV1

Linear

 μ Abs.:

30470



A= 0.0000e+000

B= 3.2761e-004

C= 5.1119e-002

Rho= 0.9998764

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	-59	0.000	-59				
0.2	0.200	0.225	0.025	532	0.0 %	532				
2.5	2.500	2.499	-0.001	7473	0.0 %	7473				
5.0	5.000	4.873	-0.127	14719	0.0 %	14719				
7.5	7.500	7.537	0.037	22849	0.0 %	22849				
10.0	10.000	10.033	0.033	30470	0.0 %	30470				

LB89301 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Aug 2017 12:17:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-59	-	0.0000	245.1	PPB	07 Aug 2017 12:20:45	S	Std
0.2 - 1	50.2	532	-	0.2000	245.1	PPB	07 Aug 2017 12:22:54	S	Std
2.5 - 1	52.5	7473	-	2.5000	245.1	PPB	07 Aug 2017 12:25:02	S	Std
5.0 - 1	55	14719	-	5.0000	245.1	PPB	07 Aug 2017 12:27:11	S	Std
7.5 - 1	57.5	22849	-	7.5000	245.1	PPB	07 Aug 2017 12:29:18	S	Std
10.0 - 1	510	30470	-	10.0000	245.1	PPB	07 Aug 2017 12:31:26	S	Std
ICV98 - 1	ICV98	11522	3.8259	-245.1	PPB	PPB	07 Aug 2017 12:34:02	U	SMPL
ICB98 - 1	ICB98	-147	0.0030	-245.1	PPB	PPB	07 Aug 2017 12:36:09	U	SMPL
CCV11 - 1	CCV11	14701	4.8673	-245.1	PPB	PPB	07 Aug 2017 12:38:16	U	SMPL
CCB11 - 1	CCB11	-121	0.0115	-245.1	PPB	PPB	07 Aug 2017 12:40:22	U	SMPL
CRA - 1	CRA	580	0.2411	-245.1	PPB	PPB	07 Aug 2017 12:42:30	U	SMPL
HighStd - 1	HighStd	30565	10.0646	-245.1	PPB	PPB	07 Aug 2017 12:44:36	U	SMPL
ChkStd - 1	ChkStd	20697	6.8317	-245.1	PPB	PPB	07 Aug 2017 12:46:43	U	SMPL
PB101289BL - 1	PBW	-231	-0.0246	-245.1	PPB	PPB	07 Aug 2017 12:48:52	U	SMPL
PB101289BS - 1	LCSW	12107	4.0175	-245.1	PPB	PPB	07 Aug 2017 12:51:01	U	SMPL
I4589-01 - 1	2AVE-05	-31	0.0410	-245.1	PPB	PPB	07 Aug 2017 12:53:08	U	SMPL
I4589-01DUP - 1	2AVE-05DUP	37	0.0632	-245.1	PPB	PPB	07 Aug 2017 12:55:17	U	SMPL
I4589-01MS - 1	2AVE-05MS	13759	4.5587	-245.1	PPB	PPB	07 Aug 2017 12:57:25	U	SMPL
I4589-01MSD - 1	2AVE-05MSD	13887	4.6007	-245.1	PPB	PPB	07 Aug 2017 12:59:34	U	SMPL
I4607-01 - 1	32220008-3	-112	0.0144	-245.1	PPB	PPB	07 Aug 2017 13:01:41	U	SMPL
CCV12 - 1	CCV12	14611	4.8379	-245.1	PPB	PPB	07 Aug 2017 13:03:47	U	SMPL
CCB12 - 1	CCB12	-119	0.0121	-245.1	PPB	PPB	07 Aug 2017 13:05:54	U	SMPL
I4589-01LX5 - 1		-7	0.0488	-245.1	PPB	PPB	07 Aug 2017 13:08:01	U	SMPL
I4589-01A - 1		13757	4.5581	-245.1	PPB	PPB	07 Aug 2017 13:12:39	U	SMPL
CCV13 - 1	CCV13	14595	4.8326	-245.1	PPB	PPB	07 Aug 2017 13:14:46	U	SMPL
CCB13 - 1	CCB13	-169	-0.0042	-245.1	PPB	PPB	07 Aug 2017 13:16:53	U	SMPL