


7471A

LB89309
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39732



A= 0.0000e+000

B= 2.5247e-004

C= 2.3677e-002

Rho= 0.9999399

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	44	0.000	44				
0.2	0.200	0.229	0.029	813	0.0 %	813				
2.5	2.500	2.456	-0.044	9636	0.0 %	9636				
5.0	5.000	4.957	-0.043	19540	0.0 %	19540				
7.5	7.500	7.468	-0.032	29486	0.0 %	29486	0			
10.0	10.000	10.055	0.055	39732	0.0 %	39732				

LB89309 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Aug 2017 15:02:07

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	44	-	0.00007471A	PPB	07 Aug 2017 15:23:52	S	Std
0.2 - 1	50.2	813	-	0.20007471A	PPB	07 Aug 2017 15:26:10	S	Std
2.5 - 1	52.5	9636	-	2.50007471A	PPB	07 Aug 2017 15:28:28	S	Std
5.0 - 1	55	19540	-	5.00007471A	PPB	07 Aug 2017 15:30:45	S	Std
7.5 - 1	57.5	29486	-	7.50007471A	PPB	07 Aug 2017 15:33:03	S	Std
10.0 - 1	510	39732	-	10.00007471A	PPB	07 Aug 2017 15:35:20	S	Std
ICV99 - 1	ICV99	15031	3.8186	-7471A	PPB	07 Aug 2017 15:38:15	U	SMPL
ICB99 - 1	ICB99	-67	0.0068	-7471A	PPB	07 Aug 2017 15:40:31	U	SMPL
CCV14 - 1	CCV14	19547	4.9587	-7471A	PPB	07 Aug 2017 15:42:46	U	SMPL
CCB14 - 1	CCB14	-165	-0.0180	-7471A	PPB	07 Aug 2017 15:45:02	U	SMPL
CRA - 1	CRA	689	0.1976	-7471A	PPB	07 Aug 2017 15:47:19	U	SMPL
HighStd - 1	HighStd	40270	10.1907	-7471A	PPB	07 Aug 2017 15:49:35	U	SMPL
ChkStd - 1	ChkStd	27811	7.0451	-7471A	PPB	07 Aug 2017 15:51:50	U	SMPL
PB101297BL - 1	PBS	-139	-0.0114	-7471A	PPB	07 Aug 2017 15:54:09	U	SMPL
PB101297BS - 1	LCSS	14125	3.5898	-7471A	PPB	07 Aug 2017 15:56:26	U	SMPL
I4623-09 - 1	TEC-S1COMP	3118	0.8109	-7471A	PPB	07 Aug 2017 15:58:43	U	SMPL
I4623-12 - 1	TEC-S2COMP	116	0.0530	-7471A	PPB	07 Aug 2017 16:01:00	U	SMPL
I4623-12DUP - 1	TEC-S2COMPDU	227	0.0810	-7471A	PPB	07 Aug 2017 16:05:52	U	SMPL
I4623-12MS - 1	TEC-S2COMPMS	14711	3.7378	-7471A	PPB	07 Aug 2017 16:08:09	U	SMPL
I4623-12MSD - 1	TEC-S2COMPMSD	14707	3.7368	-7471A	PPB	07 Aug 2017 16:10:25	U	SMPL
CCV15 - 1	CCV15	19630	4.9797	-7471A	PPB	07 Aug 2017 16:12:41	U	SMPL
CCB15 - 1	CCB15	-177	-0.0210	-7471A	PPB	07 Aug 2017 16:14:57	U	SMPL
I4625-01 - 1	CLINTON-AVE-COMP-1	901	0.2512	-7471A	PPB	07 Aug 2017 16:17:13	U	SMPL
I4625-04 - 1	CLINTON-AVE-COMP-2	1293	0.3501	-7471A	PPB	07 Aug 2017 16:19:29	U	SMPL
I4636-01 - 1	SU-03-080417	2109	0.5561	-7471A	PPB	07 Aug 2017 16:21:46	U	SMPL
I4638-10 - 1	SP-COMP	3994	1.0320	-7471A	PPB	07 Aug 2017 16:24:02	U	SMPL
I4638-12 - 1	303	6697	1.7145	-7471A	PPB	07 Aug 2017 16:26:18	U	SMPL
I4639-01 - 1	SA-06-080317-A	4078	1.0533	-7471A	PPB	07 Aug 2017 16:28:35	U	SMPL
I4639-04 - 1	SA-06-080317-B	821	0.2310	-7471A	PPB	07 Aug 2017 16:30:52	U	SMPL
I4642-01 - 1	BU-03-080417	946	0.2625	-7471A	PPB	07 Aug 2017 16:33:10	U	SMPL
I4646-01 - 1	E2-M9-BASE	10237	2.6082	-7471A	PPB	07 Aug 2017 16:35:28	U	SMPL
I4646-02 - 1	E2-M10-BASE	2107	0.5556	-7471A	PPB	07 Aug 2017 16:37:44	U	SMPL
CCV16 - 1	CCV16	19779	5.0173	-7471A	PPB	07 Aug 2017 16:40:01	U	SMPL
CCB16 - 1	CCB16	-192	-0.0248	-7471A	PPB	07 Aug 2017 16:42:17	U	SMPL
I4646-03 - 1	E2-M10-WSW	1735	0.4617	-7471A	PPB	07 Aug 2017 16:44:34	U	SMPL
I4646-04 - 1	E2-L8-BASE	147926	37.3706	-7471A	PPB	07 Aug 2017 16:46:50	U	SMPL
I4646-05 - 1	E2-L10-BASE	2156	0.5680	-7471A	PPB	07 Aug 2017 16:49:06	U	SMPL
I4646-06 - 1	E2-L10-ESW	3804	0.9841	-7471A	PPB	07 Aug 2017 16:51:26	U	SMPL
I4623-12LX5 - 1		14	0.0272	-7471A	PPB	07 Aug 2017 16:53:43	U	SMPL
I4623-12A - 1		15681	3.9827	-7471A	PPB	07 Aug 2017 16:56:00	U	SMPL
I4646-04DLX10 - 1		14882	3.7809	-7471A	PPB	07 Aug 2017 16:58:17	U	SMPL
CCV17 - 1	CCV17	19661	4.9875	-7471A	PPB	07 Aug 2017 17:00:34	U	SMPL
CCB17 - 1	CCB17	-108	-0.0036	-7471A	PPB	07 Aug 2017 17:02:52	U	SMPL