

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB89323

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP56937
calibration std. phosphate 0.5 ppm	WP56936
calibration std. phosphate 0.3 ppm	WP56935
calibration std. 0 ppm	WP56932
phosphate ICV std.	WP56939
phosphate CCV std.	WP56938
calibration std. phosphate 0.1 ppm	WP56934
calibration std. phosphate 0.05 ppm	WP56933
Combined reagent	WP57044
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Ascorbic Acid	WP57043
Sodium hydroxide, 1N	WP57036

Intercept: -0.0008Slope: 0.6506Regression: 0.99981

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	08/08/2017	13:31
2	CAL2	0.05	1	50	50	0.033	08/08/2017	13:32
3	CAL3	0.10	1	50	50	0.065	08/08/2017	13:33
4	CAL4	0.30	1	50	50	0.186	08/08/2017	13:34
5	CAL5	0.50	1	50	50	0.331	08/08/2017	13:34
6	CAL6	1.00	1	50	50	0.649	08/08/2017	13:35



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB89323

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.502	08/08/2017	13:36
2	ICB		1	50	50	0.000	0.001	08/08/2017	13:36
3	CCV1	0.50	1	50	50	0.331	0.510	08/08/2017	13:37
4	CCB1		1	50	50	0.000	0.001	08/08/2017	13:37
5	PB101343BL		1	50	50	0.000	0.001	08/08/2017	13:38
6	PB101343BS	0.50	1	50	50	0.328	0.505	08/08/2017	13:38
7	I4602-01		1	50	50	1.710	2.630	08/08/2017	13:39
8	I4602-01DUP		1	50	50	1.715	2.637	08/08/2017	13:39
9	I4602-01MS	0.50	1	50	50	2.102	3.232	08/08/2017	13:40
10	I4602-01MSD	0.50	1	50	50	2.114	3.251	08/08/2017	13:40
11	I4602-01		10	50	50	0.513	0.790	08/08/2017	13:41
12	I4602-01DUP		10	50	50	0.518	0.797	08/08/2017	13:41
13	I4602-01MS	0.50	10	50	50	0.523	0.805	08/08/2017	13:42
14	I4602-01MSD	0.50	10	50	50	0.529	0.814	08/08/2017	13:42
15	CCV2	0.50	1	50	50	0.328	0.505	08/08/2017	13:43
16	CCB2		1	50	50	0.000	0.001	08/08/2017	13:43