



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 08/08/2017DATE & TIME IN: 08/07/2017 17:00OVEN TEMP: 98 °CRun Number: LB89332DATE & TIME OUT: 08/08/2017 12:40OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-02

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB89332BL	LB89332BL	147.0283	147.0283	100	147.0283	147.0283	147.028	0	0
2	I4589-01	2AVE-05	152.9728	152.9728	100	153.1754	153.1754	153.175	0.2026	2026
3	I4589-01DUP	2AVE-05DUP	147.149	147.149	100	147.351	147.351	147.351	0.202	2020

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L =  $\frac{D}{A} \times 1000 \times 1000$

# WORKLIST(Hardcopy Internal Chain)

20389332

Worklist Name : TS080717

Worklist ID : 101596

Date : 8/7/2017 2:50:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/09/2017	Water	I4589-01	TS	Cool 4 deg C	DIST01	K62	2AVE-05	08/02/2017	SM2540 B

Date/Time

8/7/17

Received by:

DP

Relinquished by:

JS

Date/Time

8/7/17

Received by:

SL

Relinquished by:

DP