



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89345

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/08/2017	16:15
2	CAL2	1	Water			7.00	08/08/2017	16:16
3	CAL3	1	Water			10.03	08/08/2017	16:17
4	ICV	1	Water			7.01	08/08/2017	16:18
5	CCV1	1	Water			2.02	08/08/2017	16:19
6	I4666-03	1	Solid	20.02	20	8.32	08/08/2017	16:20
7	I4666-03DUP	1	Solid	20.04	20	8.33	08/08/2017	16:21
8	I4681-07	1	Solid	20.01	20	8.63	08/08/2017	16:22
9	I4681-08	1	Solid	20.04	20	8.28	08/08/2017	16:24
10	I4681-09	1	Solid	20.05	20	8.06	08/08/2017	16:25
11	CCV2	1	Water			12.01	08/08/2017	16:26

WORKLIST(Hardcopy Internal Chain)

1389345

WorkList Name : CORROSIVITY I4666

WorkList ID : 101634

Date : 8/8/2017 12:39:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2017	Solid	I4666-03	Corrosivity	Cool 4 deg C	PSEG01	L31	RO-124019-72-12014	08/07/2017	9045C
	Solid	I4681-07	Corrosivity	Cool 4 deg C	LIRO01	L42	SB-17-COMP	08/08/2017	9045C
	Solid	I4681-08	Corrosivity	Cool 4 deg C	LIRO01	L42	SB-18-COMP	08/08/2017	9045C
	Solid	I4681-09	Corrosivity	Cool 4 deg C	LIRO01	L42	SB-09-COMP	08/08/2017	9045C

Date/Time 08/08/17 12:40 PM
 Received by: JP
 Relinquished by: CD

Date/Time 08/08/17 6:00 PM
 Received by: CP
 Relinquished by: JP