



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89346

Temp.: 21.3 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/08/2017	17:50
2	CAL2	1	Water			7.00	08/08/2017	17:51
3	CAL3	1	Water			10.02	08/08/2017	17:52
4	ICV	1	Water			6.99	08/08/2017	17:53
5	CCV1	1	Water			2.01	08/08/2017	17:55
6	I4681-10	1	Water			7.09	08/08/2017	18:00
7	I4681-11	1	Water			6.82	08/08/2017	18:01
8	I4681-11DUP	1	Water			6.83	08/08/2017	18:02
9	CCV2	1	Water			12.02	08/08/2017	18:04

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14681

WorkList ID : 101646

Date : 8/8/2017 5:29:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4681-10 J	pH	Cool 4 deg C	LIRO01	L42	TWP-17	08/08/2017	SM 4500-PH B
	Water	I4681-11 Z	pH	Cool 4 deg C	LIRO01	L42	TWP-09	08/08/2017	SM 4500-PH B

Date/Time 08/08/17 5:30 PM
 Received by: JP
 Relinquished by: CR

Date/Time 08/08/17 6:30 PM
 Received by: CR
 Relinquished by: JA