


7470A

LB89351
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34490



A= 0.0000e+000

B= 2.8821e-004

C= 2.6915e-002

Rho= 0.9999228

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	66	0.000	66				
0.2	0.200	0.192	-0.008	573	0.0 %	573				
2.5	2.500	2.430	-0.070	8337	0.0 %	8337				
5.0	5.000	4.998	-0.002	17249	0.0 %	17249				
7.5	7.500	7.567	0.067	26160	0.0 %	26160				
10.0	10.000	9.967	-0.033	34490	0.0 %	34490				

LB89351 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Aug 2017 09:15:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	66	-	0.0000	7470A	PPB	09 Aug 2017 09:21:32	S	Std
0.2 - 1	SO 2	573	-	0.2000	7470A	PPB	09 Aug 2017 09:23:40	S	Std
2.5 - 1	SO 5	8337	-	2.5000	7470A	PPB	09 Aug 2017 09:25:49	S	Std
5.0 - 1	SO 5	17249	-	5.0000	7470A	PPB	09 Aug 2017 09:27:57	S	Std
7.5 - 1	SO 5	26160	-	7.5000	7470A	PPB	09 Aug 2017 09:30:05	S	Std
10.0 - 1	SO	34490	-	10.0000	7470A	PPB	09 Aug 2017 09:32:13	S	Std
ICV01 - 1	ICV01	13442	3.9011	-	7470A	PPB	09 Aug 2017 09:34:41	U	SMPL
ICB01 - 1	ICB01	-13	0.0232	-	7470A	PPB	09 Aug 2017 09:40:07	U	SMPL
CCV18 - 1	CCV18	17018	4.9317	-	7470A	PPB	09 Aug 2017 09:42:13	U	SMPL
CCB18 - 1	CCB18	-124	-0.0088	-	7470A	PPB	09 Aug 2017 09:44:20	U	SMPL
CRA - 1	CRA	555	0.1869	-	7470A	PPB	09 Aug 2017 09:46:27	U	SMPL
HighStd - 1	HighStd	35083	10.1383	-	7470A	PPB	09 Aug 2017 09:48:33	U	SMPL
ChkStd - 1	ChkStd	24447	7.0729	-	7470A	PPB	09 Aug 2017 09:50:40	U	SMPL
PB101320BL - 1	PBW	-126	-0.0094	-	7470A	PPB	09 Aug 2017 09:52:49	U	SMPL
PB101320BS - 1	LCSW	12297	3.5711	-	7470A	PPB	09 Aug 2017 09:54:57	U	SMPL
PB101320TB - 1	PB101286TB	-75	0.0053	-	7470A	PPB	09 Aug 2017 09:57:05	U	SMPL
I4625-03 - 1	CLINTON-AVE-COMP-1	0	0.0269	-	7470A	PPB	09 Aug 2017 09:59:13	U	SMPL
I4625-06 - 1	CLINTON-AVE-COMP-2	57	0.0433	-	7470A	PPB	09 Aug 2017 10:01:21	U	SMPL
I4636-03 - 1	SU-03-080417	65	0.0456	-	7470A	PPB	09 Aug 2017 10:03:30	U	SMPL
I4638-11 - 1	SP-COMP	19	0.0324	-	7470A	PPB	09 Aug 2017 10:05:37	U	SMPL
CCV19 - 1	CCV19	17253	4.9995	-	7470A	PPB	09 Aug 2017 10:07:44	U	SMPL
CCB19 - 1	CCB19	-99	-0.0016	-	7470A	PPB	09 Aug 2017 10:09:51	U	SMPL
I4638-13 - 1	303	4	0.0281	-	7470A	PPB	09 Aug 2017 10:11:58	U	SMPL
I4639-03 - 1	SA-06-080317-A	16	0.0315	-	7470A	PPB	09 Aug 2017 10:14:04	U	SMPL
I4639-06 - 1	SA-06-080317-B	57	0.0433	-	7470A	PPB	09 Aug 2017 10:16:11	U	SMPL
I4642-03 - 1	BU-03-080417	58	0.0436	-	7470A	PPB	09 Aug 2017 10:18:17	U	SMPL
I4642-03DUP - 1	BU-03-080417DUP	71	0.0474	-	7470A	PPB	09 Aug 2017 10:20:25	U	SMPL
I4642-03MS - 1	BU-03-080417MS	12435	3.6108	-	7470A	PPB	09 Aug 2017 10:22:33	U	SMPL
I4642-03MSD - 1	BU-03-080417MSD	12557	3.6460	-	7470A	PPB	09 Aug 2017 10:24:41	U	SMPL
I4642-03LX5 - 1		-149	-0.0160	-	7470A	PPB	09 Aug 2017 10:26:49	U	SMPL
I4642-03A - 1		14031	4.0708	-	7470A	PPB	09 Aug 2017 10:28:58	U	SMPL
CCV20 - 1	CCV20	17009	4.9291	-	7470A	PPB	09 Aug 2017 10:31:04	U	SMPL
CCB20 - 1	CCB20	1	0.0272	-	7470A	PPB	09 Aug 2017 10:35:19	U	SMPL