



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89352

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/09/2017	09:00
2	CAL2	1	Water			7.01	08/09/2017	09:01
3	CAL3	1	Water			10.03	08/09/2017	09:02
4	ICV	1	Water			7.00	08/09/2017	09:05
5	CCV1	1	Water			2.01	08/09/2017	09:06
6	I4346-06	1	Solid	20.02	20	5.70	08/09/2017	09:08
7	CCV2	1	Water			12.02	08/09/2017	09:10