


7471A

LB 89359
INSTRUMENT ID! CV1

Linear

 μ Abs.:

38500



A= 0.0000e+000

B= 2.5780e-004

C= 1.9548e-002

Rho= 0.9998901

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-39	0.000	-39				
0.2	0.200	0.201	0.001	704	0.0 %	704				
2.5	2.500	2.485	-0.015	9565	0.0 %	9565				
5.0	5.000	4.950	-0.050	19124	0.0 %	19124				
7.5	7.500	7.610	0.110	29443	0.0 %	29443	0			
10.0	10.000	9.945	-0.055	38500	0.0 %	38500				

LB89359 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 09 Aug 2017 11:40:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-39	-	0.00007471A	PPB		09 Aug 2017 11:42:05	S	Std
0.2 - 1	50-2	704	-	0.20007471A	PPB		09 Aug 2017 11:44:22	S	Std
2.5 - 1	52.5	9565	-	2.50007471A	PPB		09 Aug 2017 11:46:40	S	Std
5.0 - 1	55	19124	-	5.00007471A	PPB		09 Aug 2017 11:48:57	S	Std
7.5 - 1	57.5	29443	-	7.50007471A	PPB		09 Aug 2017 11:51:15	S	Std
10.0 - 1	510	38500	-	10.00007471A	PPB		09 Aug 2017 11:53:32	S	Std
ICV02 - 1	ICV02	15007	3.8883	-7471A	PPB		09 Aug 2017 11:56:59	U	SMPL
ICB02 - 1	ICB02	-92	-0.0042	-7471A	PPB		09 Aug 2017 11:59:14	U	SMPL
CCV21 - 1	CCV21	17848	4.6207	-7471A	PPB		09 Aug 2017 12:01:30	U	SMPL
CCB21 - 1	CCB21	-165	-0.0230	-7471A	PPB		09 Aug 2017 12:03:46	U	SMPL
CRA - 1	CRA	663	0.1905	-7471A	PPB		09 Aug 2017 12:06:02	U	SMPL
HighStd - 1	HighStd	38716	10.0004	-7471A	PPB		09 Aug 2017 12:08:18	U	SMPL
ChkStd - 1	ChkStd	26681	6.8978	-7471A	PPB		09 Aug 2017 12:10:34	U	SMPL
PB101371BL - 1	PBS	-233	-0.0405	-7471A	PPB		09 Aug 2017 12:12:52	U	SMPL
PB101371BS - 1	LCSS	14475	3.7511	-7471A	PPB		09 Aug 2017 12:15:09	U	SMPL
I4649-02 - 1	B-13(0-10)COMP	92846	23.9548	-7471A	PPB		09 Aug 2017 12:17:26	U	SMPL
I4649-04 - 1	B-14(0-10)COMP	12976	3.3647	-7471A	PPB		09 Aug 2017 12:19:43	U	SMPL
I4652-02 - 1	L-1338-A-COMP	329655	85.0031	-7471A	PPB		09 Aug 2017 12:22:03	U	SMPL
I4652-07 - 1	L-1338-B-COMP	181445	46.7952	-7471A	PPB		09 Aug 2017 12:24:21	U	SMPL
I4655-01 - 1	TP7-WC1	9521	2.4740	-7471A	PPB		09 Aug 2017 12:27:05	U	SMPL
CCV22 - 1	CCV22	18497	4.7880	-7471A	PPB		09 Aug 2017 12:29:31	U	SMPL
CCB22 - 1	CCB22	-261	-0.0477	-7471A	PPB		09 Aug 2017 12:31:47	U	SMPL
I4655-02 - 1	TP8N-WC1	2474	0.6573	-7471A	PPB		09 Aug 2017 12:34:03	U	SMPL
I4655-03 - 1	TP7DEL-WC1	7824	2.0365	-7471A	PPB		09 Aug 2017 12:36:19	U	SMPL
I4657-01 - 1	CL-01-080717-A	2726	0.7223	-7471A	PPB		09 Aug 2017 12:38:35	U	SMPL
I4657-04 - 1	CL-01-080717-B	7491	1.9507	-7471A	PPB		09 Aug 2017 12:40:52	U	SMPL
I4657-07 - 1	CL-01-080717-C	762	0.2160	-7471A	PPB		09 Aug 2017 12:43:09	U	SMPL
I4657-10 - 1	CL-01-080717-D	1055	0.2915	-7471A	PPB		09 Aug 2017 12:45:26	U	SMPL
I4657-13 - 1	CL-01-080717-E	1408	0.3825	-7471A	PPB		09 Aug 2017 12:47:43	U	SMPL
I4658-01 - 1	E2-L9-SSW	27414	7.0867	-7471A	PPB		09 Aug 2017 12:50:00	U	SMPL
I4658-02 - 1	E2-L9-WSW	37273	9.6284	-7471A	PPB		09 Aug 2017 12:52:17	U	SMPL
I4658-03 - 1	E2-L9-NSW	9473	2.4616	-7471A	PPB		09 Aug 2017 12:54:34	U	SMPL
CCV23 - 1	CCV23	18713	4.8437	-7471A	PPB		09 Aug 2017 12:56:52	U	SMPL
CCB23 - 1	CCB23	-184	-0.0279	-7471A	PPB		09 Aug 2017 12:59:09	U	SMPL
I4658-04 - 1	E2-L9-BASE-WEST	18521	4.7942	-7471A	PPB		09 Aug 2017 13:01:25	U	SMPL
I4658-05 - 1	E2-L9-ESW	117083	30.2030	-7471A	PPB		09 Aug 2017 13:03:42	U	SMPL
I4658-06 - 1	E2-L9-BASE	6222	1.6236	-7471A	PPB		09 Aug 2017 13:05:58	U	SMPL
I4658-06DUP - 1	E2-L9-BASEDUP	6778	1.7669	-7471A	PPB		09 Aug 2017 13:08:17	U	SMPL
I4658-06MS - 1	E2-L9-BASEMS	19753	5.1118	-7471A	PPB		09 Aug 2017 13:10:34	U	SMPL
I4658-06MSD - 1	E2-L9-BASEMSD	21376	5.5302	-7471A	PPB		09 Aug 2017 13:12:51	U	SMPL
PB101374BL - 1	PBS	-181	-0.0271	-7471A	PPB		09 Aug 2017 13:15:08	U	SMPL
PB101374BS - 1	LCSS	14592	3.7813	-7471A	PPB		09 Aug 2017 13:20:15	U	SMPL
I4650-01 - 1	SP-001-A-COMP	2928	0.7744	-7471A	PPB		09 Aug 2017 13:22:32	U	SMPL
I4650-01DUP - 1	SP-001-A-COMPDUP	3095	0.8174	-7471A	PPB		09 Aug 2017 13:27:32	U	SMPL
CCV24 - 1	CCV24	18780	4.8609	-7471A	PPB		09 Aug 2017 13:29:48	U	SMPL
CCB24 - 1	CCB24	-142	-0.0171	-7471A	PPB		09 Aug 2017 13:32:04	U	SMPL
I4650-01MS - 1	SP-001-A-COMPMS	19565	5.0633	-7471A	PPB		09 Aug 2017 13:36:56	U	SMPL
I4650-01MSD - 1	SP-001-A-COMPMSD	18347	4.7493	-7471A	PPB		09 Aug 2017 13:39:13	U	SMPL
I4666-01 - 1	RO-124019-72-12014	499	0.1482	-7471A	PPB		09 Aug 2017 13:41:29	U	SMPL
I4677-01 - 1	E2-D10-ESW	232216	59.8837	-7471A	PPB		09 Aug 2017 13:43:45	U	SMPL
I4677-02 - 1	E2-D10-WSW	146325	37.7414	-7471A	PPB		09 Aug 2017 13:46:02	U	SMPL
I4677-03 - 1	E2-D10-BASE	28492	7.3647	-7471A	PPB		09 Aug 2017 13:48:33	U	SMPL
I4677-04 - 1	E2-E10-ESW	189030	48.7506	-7471A	PPB		09 Aug 2017 13:50:55	U	SMPL
I4677-05 - 1	E2-E10-WSW	1218845	HIGH	-7471A	PPB		09 Aug 2017 13:53:14	X	
I4677-06 - 1	E2-E10-BASE	15872	4.1113	-7471A	PPB		09 Aug 2017 13:55:16	U	SMPL
I4677-07 - 1	E2-F9-WSW	21134	5.4678	-7471A	PPB		09 Aug 2017 13:59:06	U	SMPL
CCV25 - 1	CCV25	18384	4.7589	-7471A	PPB		09 Aug 2017 14:01:23	U	SMPL
CCB25 - 1	CCB25	-245	-0.0436	-7471A	PPB		09 Aug 2017 14:06:21	U	SMPL
I4677-08 - 1	E2-F9-BASE	27351	7.0705	-7471A	PPB		09 Aug 2017 14:08:37	U	SMPL
I4677-09 - 1	E2-F10-BASE	10843	2.8148	-7471A	PPB		09 Aug 2017 14:10:53	U	SMPL
I4677-10 - 1	E2-F10-ESW	254838	65.7156	-7471A	PPB		09 Aug 2017 14:13:11	U	SMPL
I4658-06LX5 - 1		1017	0.2817	-7471A	PPB		09 Aug 2017 14:18:29	U	SMPL
I4658-06A - 1		23000	5.9488	-7471A	PPB		09 Aug 2017 14:20:45	U	SMPL
I4650-01LX5 - 1		399	0.1224	-7471A	PPB		09 Aug 2017 14:23:02	U	SMPL
I4650-01A - 1		18574	4.8078	-7471A	PPB		09 Aug 2017 14:25:19	U	SMPL
CCV26 - 1	CCV26	18004	4.6609	-7471A	PPB		09 Aug 2017 14:29:29	U	SMPL
CCB26 - 1	CCB26	-349	-0.0704	-7471A	PPB		09 Aug 2017 14:31:56	U	SMPL
I4649-02DLX5 - 1		18700	4.8403	-7471A	PPB		09 Aug 2017 14:34:13	U	SMPL
I4652-02DLX10 - 1		35509	9.1736	-7471A	PPB		09 Aug 2017 14:36:30	U	SMPL
I4652-07DLX10 - 1		18114	4.6893	-7471A	PPB		09 Aug 2017 14:38:47	U	SMPL
I4658-05DLX5 - 1		22496	5.8189	-7471A	PPB		09 Aug 2017 14:41:05	U	SMPL
I4677-01DLX10 - 1		23785	6.1512	-7471A	PPB		09 Aug 2017 14:43:22	U	SMPL
I4677-02DLX10 - 1		14571	3.7759	-7471A	PPB		09 Aug 2017 14:45:39	U	SMPL
I4677-04DLX10 - 1		18891	4.8896	-7471A	PPB		09 Aug 2017 14:47:57	U	SMPL
I4677-05DLX100 - 1		57097	14.7389	-7471A	PPB		09 Aug 2017 14:50:14	U	SMPL
I4677-10DLX10 - 1		25957	6.7111	-7471A	PPB		09 Aug 2017 14:52:31	U	SMPL
I4677-05DL2X500 - 1		11412	2.9615	-7471A	PPB		09 Aug 2017 14:54:51	U	SMPL
CCV27 - 1	CCV27	18851	4.8792	-7471A	PPB		09 Aug 2017 14:57:09	U	SMPL
CCB27 - 1	CCB27	-184	-0.0279	-7471A	PPB		09 Aug 2017 14:59:26	U	SMPL