

7470A


LB89368

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36513



A= 0.0000e+000

B= 2.7434e-004

C= 5.5647e-002

Rho= 0.9998562

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.063	0.063	27	0.000	27				
0.2	0.200	0.222	0.022	606	0.0 %	606				
2.5	2.500	2.411	-0.089	8584	0.0 %	8584				
5.0	5.000	5.008	0.008	18051	0.0 %	18051				
7.5	7.500	7.424	-0.076	26860	0.0 %	26860				
10.0	10.000	10.072	0.072	36513	0.0 %	36513				

LB89368 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Aug 2017 15:45:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.00007470A	PPB		09 Aug 2017 16:01:38	S	Std
0.2 - 1	50.2	606	-	0.20007470A	PPB		09 Aug 2017 16:03:47	S	Std
2.5 - 1	52.5	8584	-	2.50007470A	PPB		09 Aug 2017 16:05:55	S	Std
5.0 - 1	55	18051	-	5.00007470A	PPB		09 Aug 2017 16:08:03	S	Std
7.5 - 1	57.5	26860	-	7.50007470A	PPB		09 Aug 2017 16:12:45	S	Std
10.0 - 1	51	36513	-	10.00007470A	PPB		09 Aug 2017 16:17:21	S	Std
ICV03 - 1	ICV03	13351	3.7183	-7470A	PPB		09 Aug 2017 16:19:50	U	SMPL
ICB03 - 1	ICB03	-119	0.0230	-7470A	PPB		09 Aug 2017 16:21:57	U	SMPL
CCV28 - 1	CCV28	17640	4.8949	-7470A	PPB		09 Aug 2017 16:24:03	U	SMPL
CCB28 - 1	CCB28	-101	0.0279	-7470A	PPB		09 Aug 2017 16:26:10	U	SMPL
CRA - 1	CRA	580	0.2148	-7470A	PPB		09 Aug 2017 16:28:16	U	SMPL
HighStd - 1	HighStd	36481	10.0637	-7470A	PPB		09 Aug 2017 16:30:23	U	SMPL
ChkStd - 1	ChkStd	24107	6.6691	-7470A	PPB		09 Aug 2017 16:32:30	U	SMPL
PB101384BL - 1	PBW	-158	0.0123	-7470A	PPB		09 Aug 2017 16:34:38	U	SMPL
PB101384BS - 1	LCSW	13108	3.6516	-7470A	PPB		09 Aug 2017 16:36:47	U	SMPL
PB101384TB - 1	PB101322TB	-95	0.0296	-7470A	PPB		09 Aug 2017 16:38:54	U	SMPL
I4649-05 - 1	B-13(0-10)COMP	-3	0.0548	-7470A	PPB		09 Aug 2017 16:41:02	U	SMPL
I4649-06 - 1	B-14(0-10)COMP	8	0.0578	-7470A	PPB		09 Aug 2017 16:43:11	U	SMPL
I4650-03 - 1	SP-001-A-COMP	-18	0.0507	-7470A	PPB		09 Aug 2017 16:45:19	U	SMPL
I4652-06 - 1	L-1338-A-COMP	20	0.0611	-7470A	PPB		09 Aug 2017 16:47:25	U	SMPL
CCV29 - 1	CCV29	17282	4.7967	-7470A	PPB		09 Aug 2017 16:49:32	U	SMPL
CCB29 - 1	CCB29	-89	0.0312	-7470A	PPB		09 Aug 2017 16:51:39	U	SMPL
I4652-11 - 1	L-1338-B-COMP	-7	0.0537	-7470A	PPB		09 Aug 2017 16:53:46	U	SMPL
I4655-04 - 1	TP7-WC1	13	0.0592	-7470A	PPB		09 Aug 2017 16:55:53	U	SMPL
I4655-05 - 1	TP8N-WC1	12	0.0589	-7470A	PPB		09 Aug 2017 16:58:00	U	SMPL
I4655-06 - 1	TP7DEL-WC1	94	0.0814	-7470A	PPB		09 Aug 2017 17:00:07	U	SMPL
I4657-03 - 1	CL-01-080717-A	-16	0.0513	-7470A	PPB		09 Aug 2017 17:02:15	U	SMPL
I4657-06 - 1	CL-01-080717-B	0	0.0556	-7470A	PPB		09 Aug 2017 17:04:22	U	SMPL
I4657-09 - 1	CL-01-080717-C	9	0.0581	-7470A	PPB		09 Aug 2017 17:06:31	U	SMPL
I4657-12 - 1	CL-01-080717-D	11	0.0587	-7470A	PPB		09 Aug 2017 17:08:39	U	SMPL
I4657-15 - 1	CL-01-080717-E	-1	0.0554	-7470A	PPB		09 Aug 2017 17:10:48	U	SMPL
I4657-15DUP - 1	CL-01-080717-EDUP	4	0.0567	-7470A	PPB		09 Aug 2017 17:12:55	U	SMPL
CCV30 - 1	CCV30	17324	4.8082	-7470A	PPB		09 Aug 2017 17:15:02	U	SMPL
CCB30 - 1	CCB30	-91	0.0307	-7470A	PPB		09 Aug 2017 17:17:08	U	SMPL
I4657-15MS - 1	CL-01-080717-EMS	11876	3.3137	-7470A	PPB		09 Aug 2017 17:19:16	U	SMPL
I4657-15MSD - 1	CL-01-080717-EMSD	12376	3.4508	-7470A	PPB		09 Aug 2017 17:21:23	U	SMPL
PB101390BL - 1	PBW	-80	0.0337	-7470A	PPB		09 Aug 2017 17:23:30	U	SMPL
PB101390BS - 1	LCSW	12765	3.5576	-7470A	PPB		09 Aug 2017 17:25:37	U	SMPL
PB101390TB - 1	PB101349TB	-81	0.0334	-7470A	PPB		09 Aug 2017 17:27:44	U	SMPL
I4666-03 - 1	RO-124019-72-12014	-3	0.0548	-7470A	PPB		09 Aug 2017 17:29:51	U	SMPL
I4681-07 - 1	SB-17-COMP	-6	0.0540	-7470A	PPB		09 Aug 2017 17:32:00	U	SMPL
I4681-08 - 1	SB-18-COMP	6	0.0573	-7470A	PPB		09 Aug 2017 17:34:08	U	SMPL
I4681-09 - 1	SB-09-COMP	-8	0.0535	-7470A	PPB		09 Aug 2017 17:36:17	U	SMPL
I4681-09DUP - 1	SB-09-COMPDUP	4	0.0567	-7470A	PPB		09 Aug 2017 17:38:23	U	SMPL
CCV31 - 1	CCV31	17525	4.8634	-7470A	PPB		09 Aug 2017 17:40:31	U	SMPL
CCB31 - 1	CCB31	-88	0.0315	-7470A	PPB		09 Aug 2017 17:42:38	U	SMPL
I4681-09MS - 1	SB-09-COMPMS	12573	3.5049	-7470A	PPB		09 Aug 2017 17:44:46	U	SMPL
I4681-09MSD - 1	SB-09-COMPMSD	12284	3.4256	-7470A	PPB		09 Aug 2017 17:46:53	U	SMPL
I4657-15LX5 - 1		-79	0.0340	-7470A	PPB		09 Aug 2017 17:49:00	U	SMPL
I4657-15A - 1		14117	3.9285	-7470A	PPB		09 Aug 2017 17:51:07	U	SMPL
I4681-09LX5 - 1		-90	0.0310	-7470A	PPB		09 Aug 2017 17:53:15	U	SMPL
I4681-09A - 1		13464	3.7493	-7470A	PPB		09 Aug 2017 17:55:22	U	SMPL
CCV32 - 1	CCV32	17378	4.8231	-7470A	PPB		09 Aug 2017 17:57:31	U	SMPL
CCB32 - 1	CCB32	-103	0.0274	-7470A	PPB		09 Aug 2017 17:59:39	U	SMPL