



Analytical Summary Report

Analysis Method: SM4500-H B

Analyst By : gorge

Parameter: Field pH

Supervisor Review By : apatel

Run Number: LB89394

Temp.: 46.65 °C

SLOPE.: 1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
PH 10.01 BUFFER,COLOR CD 475ML	W2154

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 07 Control Limits[+/- 0.1].

True Value of CCV2 = 07 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4	08/01/2017	14:02
2	CAL2	1	Water			7	08/01/2017	14:06
3	CAL3	1	Water			10.01	08/01/2017	14:08
4	ICV	1	Water			7	08/01/2017	14:10
5	CCV1	1	Water			7	08/01/2017	14:14
6	I4563-01	1	Water			8.89	08/01/2017	15:00
7	CCV2	1	Water			7	08/01/2017	15:04