



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB89400

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	08/10/2017	09:45
2	I4681-10		1	>100.0	>212.0	08/10/2017	10:15
3	I4681-10DUP		1	>100.0	>212.0	08/10/2017	10:50
4	I4681-11		1	>100.0	>212.0	08/10/2017	11:12

Result = (Celsius * 1.8) + 32

AB89400

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP81017

WorkList ID : 101711

Date : 8/10/2017 8:54:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4681-10	Flash Point	Cool 4 deg C	LIRO01	L42	TWP-17	08/08/2017	1010A
	Water	I4681-11	Flash Point	Cool 4 deg C	LIRO01	L42	TWP-09	08/08/2017	1010A

Date/Time 8/10/17 9:00 AM
Received by: KS
Relinquished by: CP

Date/Time 8/10/17 11:15 AM
Received by: CP
Relinquished by: KS